

**LOT** PX 437  
Rev 2


**CONTROL**

 (Exp.) 2022-11-05  
(YYYY - MM - DD)

		ABX Lysebio															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA					
		XL80	XLR				XL80	XLR				XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.2	3.2			± 0.4	8.5	8.5			± 1.0	18.4	18.4			± 2.2	
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.29	2.29			± 0.16	4.59	4.59			± 0.20	5.14	5.14			± 0.25	
HB HGB	g/dl	6.0	6.0			± 0.4	13.6	13.6			± 0.5	16.8	16.8			± 0.6	
	g/l	60	60			± 4	136	136			± 5	168	168			± 6	
	mmol/l	3.73	3.73			± 0.25	8.45	8.45			± 0.31	10.43	10.43			± 0.37	
HT HCT	%	17.9	17.9			± 1.5	39.5	39.5			± 2.0	47.8	47.8			± 2.5	
	l/l	0.179	0.179			± 0.015	0.395	0.395			± 0.020	0.478	0.478			± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	78	78			± 5	86	86			± 5	93	93			± 5	
TGMH MCH	pg	26.2	26.2			± 2.0	29.6	29.6			± 2.0	32.7	32.7			± 2.5	
	fmol	1.63	1.63			± 0.12	1.84	1.84			± 0.12	2.03	2.03			± 0.16	
CCMH MCHC	g/dl	33.6	33.6			± 3.0	34.5	34.5			± 3.0	35.1	35.1			± 3.0	
	g/l	336	336			± 30	345	345			± 30	351	351			± 30	
	mmol/l	20.86	20.86			± 1.86	21.40	21.40			± 1.86	21.83	21.83			± 1.86	
IDR RDW	%	15.0	15.0			± 4.0	13.0	13.0			± 4.0	12.0	12.0			± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	63	63			± 20	220	220			± 30	440	440			± 50	
VPM MPV	µm <sup>3</sup> ; fl	10.4	10.4			± 2.0	9.9	9.9			± 2.0	11.0	11.0			± 2.0	
NEUT	#	1.64	1.64			± 0.35	4.61	4.61			± 0.90	13.58	13.58			± 1.90	
	%	51.1	51.1			± 10.0	54.2	54.2			± 10.0	73.8	73.8			± 10.0	
LYMPHO	#	1.21	1.21			± 0.33	3.17	3.17			± 0.70	2.87	2.87			± 1.50	
	%	37.7	37.7			± 12.0	37.3	37.3			± 8.0	15.6	15.6			± 8.0	
MONO	#	0.08	0.08			± 0.08	0.17	0.17			± 0.17	0.35	0.35			± 0.35	
	%	2.4	2.4			± 2.4	2.0	2.0			± 2.0	1.9	1.9			± 1.9	
EOS	#	0.19	0.19			± 0.19	0.30	0.30			± 0.30	0.83	0.83			± 0.83	
	%	6.0	6.0			± 6.0	3.5	3.5			± 3.5	4.5	4.5			± 4.5	
BASO	#	0.09	0.09			± 0.09	0.26	0.26			± 0.26	0.77	0.77			± 0.77	
	%	2.8	2.8			± 2.8	3.0	3.0			± 3.0	4.2	4.2			± 4.2	

**LOT** PX 437  
Rev 2

**CONTROL**

 (Exp.) 2022-11-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.2	3.2	3.2		± 0.4	8.5	8.6	8.6		± 1.0	18.9	18.8	18.4		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.37	2.36	2.35		± 0.16	4.61	4.61	4.60		± 0.20	5.15	5.15	5.16		± 0.25			
HB HGB	g/dl	6.1	6.0	6.0		± 0.4	13.7	13.6	13.6		± 0.5	16.9	16.9	16.8		± 0.6			
	g/l	61	60	60		± 4	137	136	136		± 5	169	169	168		± 6			
	mmol/l	3.79	3.73	3.73		± 0.25	8.51	8.45	8.45		± 0.31	10.49	10.49	10.43		± 0.37			
HT HCT	%	17.8	17.7	17.4		± 1.5	39.2	38.7	38.4		± 2.0	47.9	47.4	47.2		± 2.5			
	l/l	0.178	0.177	0.174		± 0.015	0.392	0.387	0.384		± 0.020	0.479	0.474	0.472		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	75	75	74.0		± 5	85	84	83.5		± 5	93	92	91.5		± 5			
TGMH MCH	pg	25.7	25.4	25.5		± 2.0	29.7	29.5	29.6		± 2.0	32.8	32.8	32.6		± 2.5			
	fmol	1.60	1.58	1.59		± 0.12	1.85	1.83	1.84		± 0.12	2.04	2.04	2.02		± 0.16			
CCMH MCHC	g/dl	34.3	33.9	34.5		± 3.0	35.0	35.1	35.4		± 3.0	35.3	35.7	35.6		± 3.0			
	g/l	343	339	345		± 30	350	351	354		± 30	353	357	356		± 30			
	mmol/l	21.31	21.05	21.43		± 1.86	21.71	21.81	21.99		± 1.86	21.91	22.15	22.10		± 1.86			
IDR RDW	%	15.0	14.5	14.0		± 4.0	13.0	12.5	12.0		± 4.0	12.0	11.5	11.0		± 4.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	68	64	66		± 20	228	226	216		± 30	457	463	438		± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.8	10.0	9.3		± 2.0	9.3	9.3	8.7		± 2.0	10.1	10.1	9.3		± 2.0			
NEUT	#	1.64	1.63	1.70		± 0.35	4.53	4.57	4.69		± 0.90	13.82	13.82	13.71		± 1.90			
	%	51.2	50.8	53.0		± 10.0	53.3	53.1	54.5		± 10.0	73.1	73.5	74.5		± 10.0			
LYMPHO	#	1.22	1.17	1.18		± 0.33	3.24	3.23	3.23		± 0.70	3.02	2.97	2.85		± 1.50			
	%	38.0	36.7	37.0		± 12.0	38.1	37.6	37.5		± 8.0	16.0	15.8	15.5		± 8.0			
MONO	#	0.09	0.09	0.08		± 0.08	0.19	0.17	0.17		± 0.17	0.40	0.34	0.28		± 0.28			
	%	2.8	2.7	2.5		± 2.5	2.2	2.0	2.0		± 2.0	2.1	1.8	1.5		± 1.5			
EOS	#	0.17	0.22	0.14		± 0.14	0.29	0.37	0.26		± 0.26	0.87	0.86	0.83		± 0.83			
	%	5.2	7.0	4.5		± 4.5	3.4	4.3	3.0		± 3.0	4.6	4.6	4.5		± 4.5			
BASO	#	0.09	0.09	0.10		± 0.09	0.26	0.26	0.26		± 0.26	0.79	0.81	0.74		± 0.74			
	%	2.8	2.8	3.0		± 2.8	3.0	3.0	3.0		± 3.0	4.2	4.3	4.0		± 4.0			

1300090306 FRONT / RECTO Ref: TEMP-0821 Rev.47