


LOT PX 438
Rev 1


CONTROL

 (Exp.) 2023-01-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.1	3.1			± 0.4	8.5	8.5			± 1.0	18.1	18.1			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.30	2.30			± 0.16	4.58	4.58			± 0.20	5.13	5.13			± 0.25			
HB HGB	g/dl	6.3	6.3			± 0.4	14.2	14.2			± 0.5	16.5	16.5			± 0.6			
	g/l	63	63			± 4	142	142			± 5	165	165			± 6			
	mmol/l	3.91	3.91			± 0.25	8.82	8.82			± 0.31	10.25	10.25			± 0.37			
HT HCT	%	18.6	18.6			± 1.5	40.8	40.8			± 2.0	46.2	46.2			± 2.5			
	l/l	0.186	0.186			± 0.015	0.408	0.408			± 0.020	0.462	0.462			± 0.025			
VGM MCV	µm ³ ; fl	81	81			± 5	89	89			± 5	90	90			± 5			
TGMH MCH	pg	27.4	27.4			± 2.0	31.0	31.0			± 2.0	32.2	32.2			± 2.5			
	fmol	1.70	1.70			± 0.12	1.93	1.93			± 0.12	2.00	2.00			± 0.16			
CCMH MCHC	g/dl	33.8	33.8			± 3.0	34.8	34.8			± 3.0	35.7	35.7			± 3.0			
	g/l	338	338			± 30	348	348			± 30	357	357			± 30			
	mmol/l	21.00	21.00			± 1.86	21.63	21.63			± 1.86	22.19	22.19			± 1.86			
IDR RDW	%	14.5	14.5			± 4.0	13.0	13.0			± 4.0	12.5	12.5			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	67	67			± 20	236	236			± 30	500	500			± 50			
VPM MPV	µm ³ ; fl	9.0	9.0			± 2.0	9.2	9.2			± 2.0	8.3	8.3			± 2.0			
NEUT	#	1.55	1.55			± 0.35	4.47	4.47			± 0.90	12.81	12.81			± 1.90			
	%	50.0	50.0			± 10.0	52.6	52.6			± 10.0	70.8	70.8			± 10.0			
LYMPHO	#	1.15	1.15			± 0.33	3.30	3.30			± 0.70	3.48	3.48			± 1.50			
	%	37.2	37.2			± 12.0	38.8	38.8			± 8.0	19.2	19.2			± 8.0			
MONO	#	0.10	0.10			± 0.10	0.24	0.24			± 0.24	0.36	0.36			± 0.36			
	%	3.3	3.3			± 3.3	2.8	2.8			± 2.8	2.0	2.0			± 2.0			
EOS	#	0.21	0.21			± 0.21	0.25	0.25			± 0.25	0.72	0.72			± 0.72			
	%	6.7	6.7			± 6.7	2.9	2.9			± 2.9	4.0	4.0			± 4.0			
BASO	#	0.09	0.09			± 0.09	0.25	0.25			± 0.25	0.72	0.72			± 0.72			
	%	2.8	2.8			± 2.8	2.9	2.9			± 2.9	4.0	4.0			± 4.0			

LOT PX 438
Rev 1

CONTROL

 (Exp.) 2023-01-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0	3.1		± 0.4	8.5	8.6	8.4		± 1.0	18.6	18.5	18.4		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.37	2.34	2.33		± 0.16	4.62	4.58	4.56		± 0.20	5.13	5.09	5.07		± 0.25			
HB HGB	g/dl	6.3	6.2	6.3		± 0.4	14.2	14.1	14.1		± 0.5	16.6	16.4	16.4		± 0.6			
	g/l	63	62	63		± 4	142	141	141		± 5	166	164	164		± 6			
HT HCT	mmol/l	3.91	3.85	3.91		± 0.25	8.82	8.76	8.76		± 0.31	10.31	10.18	10.18		± 0.37			
	%	18.7	18.3	17.9		± 1.5	41.1	40.3	40.1		± 2.0	47.2	46.3	46.1		± 2.5			
VGM MCV	l/l	0.187	0.183	0.179		± 0.015	0.411	0.403	0.401		± 0.020	0.472	0.463	0.461		± 0.025			
	µm ³ /fl	79	78	77.0		± 5	89	88	88.0		± 5	92	91	91.0		± 5			
TGMH MCH	pg	26.6	26.5	27.0		± 2.0	30.7	30.8	30.9		± 2.0	32.4	32.2	32.3		± 2.5			
	fmol	1.65	1.65	1.68		± 0.12	1.91	1.91	1.92		± 0.12	2.01	2.00	2.01		± 0.16			
CCMH MCHC	g/dl	33.6	34.0	35.1		± 3.0	34.5	35.0	35.1		± 3.0	35.2	35.4	35.5		± 3.0			
	g/l	336	340	351		± 30	345	350	351		± 30	352	354	355		± 30			
IDR RDW	mmol/l	20.90	21.09	21.81		± 1.86	21.45	21.73	21.82		± 1.86	21.84	21.99	22.07		± 1.86			
	%	15.5	14.5	14.0		± 4.0	12.5	12.0	11.5		± 4.0	12.0	11.5	11.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	69	65	67		± 20	240	234	230		± 30	493	489	480		± 50			
VPM MPV	µm ³ /fl	8.4	8.8	8.3		± 2.0	8.6	8.7	8.4		± 2.0	7.8	7.8	7.5		± 2.0			
NEUT	#	1.48	1.46	1.57		± 0.35	4.35	4.41	4.37		± 0.90	13.02	13.02	13.06		± 1.90			
	%	49.3	48.7	50.5		± 10.0	51.2	51.3	52.0		± 10.0	70.0	70.4	71.0		± 10.0			
LYMPHO	#	1.13	1.13	1.16		± 0.33	3.40	3.41	3.36		± 0.70	3.68	3.63	3.59		± 1.50			
	%	37.8	37.8	37.5		± 12.0	40.0	39.7	40.0		± 8.0	19.8	19.6	19.5		± 8.0			
MONO	#	0.12	0.12	0.11		± 0.11	0.26	0.26	0.21		± 0.21	0.41	0.37	0.28		± 0.28			
	%	4.0	3.9	3.5		± 3.5	3.1	3.0	2.5		± 2.5	2.2	2.0	1.5		± 1.5			
EOS	#	0.18	0.20	0.17		± 0.17	0.24	0.27	0.21		± 0.21	0.74	0.74	0.74		± 0.74			
	%	6.1	6.7	5.5		± 5.5	2.8	3.1	2.5		± 2.5	4.0	4.0	4.0		± 4.0			
BASO	#	0.08	0.09	0.09		± 0.08	0.25	0.25	0.25		± 0.25	0.74	0.74	0.74		± 0.74			
	%	2.8	2.9	3.0		± 2.8	2.9	2.9	3.0		± 2.9	4.0	4.0	4.0		± 4.0			

1300090306
FRONT / RECTO
Ref: TEMP-0821 Rev.47