

ABX Minotrol Retic



LOT RX 436
Rev 1

CONTROL

(Exp.) 2022-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	± 0.8	5.5	5.5	± 2.2	10.4	10.4	± 3.1			
RETIC	%	1.63	1.63	± 0.82	5.02	5.02	± 1.26	11.45	11.45	± 2.29			
RETIC	10 ⁶ /µl	0.056	0.056	± 0.028	0.174	0.174	± 0.044	0.395	0.395	± 0.079			
MFI MRI *	%	30.3	30.3	± 8.0	44.9	44.9	± 10.0	52.5	52.5	± 12.0			
GR RBC	10 ⁶ /µl	3.45	3.45	± 0.30	3.46	3.46	± 0.30	3.45	3.45	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.42 FRONT / RECTO 1300090305

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 436
Rev 1

CONTROL

(Exp.) 2022-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.5			± 0.8	5.5			± 2.2	10.4			± 3.1
RETIC	%	1.70			± 0.85	5.26			± 1.32	11.99			± 2.40
RETIC	10 ⁶ /µl	0.058			± 0.029	0.180			± 0.045	0.408			± 0.082
MFI MRI *	%	29.3			± 8.0	41.7			± 10.0	46.7			± 12.0
GR RBC	10 ⁶ /µl	3.43			± 0.30	3.43			± 0.30	3.40			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.42 BACK / VERSO 1300090305

* For research use only in Canada and USA