

ABX Minotrol Retic



LOT RX 437
Rev 1

CONTROL

(Exp.) 2022-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.3	1.3	± 0.7	5.4	5.4	± 2.2	10.7	10.7	± 3.2			
RETIC	%	1.70	1.70	± 0.85	5.38	5.38	± 1.35	11.00	11.00	± 2.20			
RETIC	10 ⁶ /µl	0.059	0.059	± 0.030	0.187	0.187	± 0.047	0.386	0.386	± 0.077			
MFI MRI *	%	30.1	30.1	± 8.0	40.3	40.3	± 10.0	44.2	44.2	± 12.0			
GR RBC	10 ⁶ /µl	3.47	3.47	± 0.30	3.48	3.48	± 0.30	3.51	3.51	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.42 FRONT / RECTO 1300090305

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 437
Rev 1

CONTROL

(Exp.) 2022-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	5.4			± 2.2	10.7			± 3.2
RETIC	%	1.83			± 0.92	5.50			± 1.38	11.04			± 2.21
RETIC	10 ⁶ /µl	0.062			± 0.031	0.187			± 0.047	0.376			± 0.075
MFI MRI *	%	28.2			± 8.0	37.9			± 10.0	40.9			± 12.0
GR RBC	10 ⁶ /µl	3.38			± 0.30	3.40			± 0.30	3.41			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.42 BACK / VERSO 1300090305

* For research use only in Canada and USA