

ABX Minotrol CRP



LOT

MC 426

CONTROL



(Exp.) 2023-04-05
(YYYY - MM - DD)

Rev 1

		ABX Lyse															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.4	7.5			± 1.0	19.7	20.4			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.34	2.33			± 0.15	4.48	4.47			± 0.18	5.48	5.46			± 0.25	
HB HGB	g/dl	6.2	6.2			± 0.4	13.2	13.4			± 0.5	17.4	17.6			± 0.7	
	g/l	62	62			± 4	132	134			± 5	174	176			± 7	
	mmol/l	3.85	3.85			± 0.25	8.20	8.32			± 0.31	10.81	10.93			± 0.44	
HT HCT	%	16.8	16.8			± 2.0	36.3	36.7			± 2.5	49.3	49.1			± 3.0	
	l/l	0.168	0.168			± 0.020	0.363	0.367			± 0.025	0.493	0.491			± 0.030	
VGM MCV	µm ³ ; fl	72	72			± 4	81	82			± 4	90	90			± 4	
TGMH MCH	pg	26.5	26.6			± 2.4	29.5	30.0			± 2.4	31.8	32.2			± 2.8	
	fmol	1.65	1.65			± 0.15	1.83	1.86			± 0.15	1.97	2.00			± 0.18	
CCMH MCHC	g/dl	36.8	37.0			± 3.0	36.4	36.6			± 3.0	35.3	35.8			± 3.0	
	g/l	368	370			± 30	364	366			± 30	353	358			± 30	
	mmol/l	22.85	22.95			± 1.86	22.59	22.70			± 1.86	21.91	22.24			± 1.86	
IDR RDW	%	14.0	14.0			± 3.0	13.5	13.0			± 3.0	13.0	12.5			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	77	76			± 20	270	269			± 40	520	514			± 60	
VPM MPV	µm ³ ; fl	8.8	7.8			± 2.0	8.4	7.7			± 2.0	8.3	7.7			± 2.0	
LYMPHO	#	1.3	1.3			± 0.2	2.5	2.5			± 0.5	3.3	3.3			± 1.1	
	%	66.0	64.0			± 8.0	34.0	33.5			± 6.0	17.0	16.0			± 5.0	
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.2	1.1			± 1.0	
	%	9.0	8.0			± 7.0	7.5	7.0			± 5.0	6.0	5.5			± 5.0	
GRANULO	#	0.5	0.6			± 0.2	4.3	4.5			± 0.5	15.2	16.0			± 1.5	
	%	25.0	28.0			± 7.0	58.5	59.5			± 6.0	77.0	78.5			± 7.0	
CRP	mg/dl	0.40	N/A			± 0.30	1.93	N/A			± 0.70	4.38	N/A			± 1.00	
	mg/l	4.0	N/A			± 3.0	19.3	N/A			± 7.0	43.8	N/A			± 10.0	

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170