

ABX Minocal

LOT

CX 478

CAL



(Exp.) 2023-02-05

Rev **2**

(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|--------------|--------------|--------|--------|-----------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 | CRP 200 | | | 60 8P 60 18P | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 8.9 | 9.0 | | | 9.0 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.54 | 4.60 | | | 4.53 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.6 | 13.4 | | | 13.7 | | | | | ± 0.2 |
| | g/l | 136 | 134 | | | 137 | | | | | ± 2 |
| | mmol/l | 8.45 | 8.34 | | | 8.51 | | | | | ± 0.12 |
| HT HCT | % | 38.1 | 38.9 | | | 39.2 | | | | | ± 1.0 |
| | l/l | 0.381 | 0.389 | | | 0.392 | | | | | ± 0.010 |
| PLA PLT | 10 ³ /mm ³ ;10 ⁹ /l | 227 | 242 | | | 241 | | | | | ± 10 |
| VMP MPV | µm ³ ; fl | 8.7 | 8.3 | | | 9.1 | | | | | ± 0.5 |

ABX Minocal

LOT

CX 478

CAL



(Exp.) 2023-02-05
(YYYY-MM-DD)

Rev **2**

| PARAMETRES PARAMETERS | UNITES UNITS | LyseVet | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|-----------------------------|--------|--|--|--|--|--|-------------------------|
| | | MICROS | | MICROS | | | | | | |
| | | Scil Vet ABC | Scil Vet ABC Plus ESV 60 | | | | | | | |
| GB WBC | 10 ⁹ /mm ³ ;10 ⁹ /l | 8.9 | 8.8 | | | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.54 | 4.53 | | | | | | | ± 0.06 |
| HB HGB | g/dl | 13.2 | 13.4 | | | | | | | ± 0.2 |
| | g/l | 132 | 134 | | | | | | | ± 2 |
| | mmol/l | 8.20 | 8.32 | | | | | | | ± 0.12 |
| HT HCT | % | 37.7 | 38.5 | | | | | | | ± 1.0 |
| | l/l | 0.377 | 0.385 | | | | | | | ± 0.010 |
| PLA PLT | 10 ³ /mm ³ ;10 ⁹ /l | 219 | 245 | | | | | | | ± 10 |
| VMP MPV | µm ³ ; fl | N/A | 9.2 | | | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.48 FRONT / RECTO 1300105208