

HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée
 B.P. 7290
 34184 Montpellier Cedex 4
 FRANCE
 www.horiba.com



LOT 22067 (rev. 01)



2 2 0 6 7

SAP 1300029568



1300029568

REF : 1300023947



2024/05/31

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400 (P400)	Pentra C400 (PC400)	Pentra C200 (PC200)	Unité	Valeur titrée	Intervalle de confiance	
				Unit	Assigned value	Confidence range	
				Einheit	Deklariertes Wert	Vertrauensintervalle	
				Unidad	Valor asignado	Intervalos de confianza	
				Unità	Valori assegnati	Livelli di fiducia	
				Unidade	Valor determinado	Intervalo de confiança	
				Enhed	Tildelt værdi	Konfidensområde	
				Enhet	Tildelat värde	Konfidensintervall	
				Μονάδα	Αναγραφόμενη τιμή	Εύρος εμπιστοσύνης	
				Jednostka	Wartość przypisana	Zakres ufności	
				Enhet	Tildelt verd	Konfidensspekter	
C				U	\oplus	Confidence range*	1 SD
AlbT2-U ABX Pentra Micro ALBUMIN 2 CP (1300032563)	•			mg/L mg/dL	65.0 6.50	52.0 - 78.0 5.20 - 7.80	6.5 0.65
		•		mg/L mg/dL	65.0 6.50	52.0 - 78.0 5.20 - 7.80	6.5 0.65
			•	mg/L mg/dL	60.7 6.07	48.5 - 72.9 4.85 - 7.29	6.1 0.61
AMY (EPS-G7) ABX Pentra Amylase CP (A11A01628)	•			U/L μkat/L	249.3 4.163	199.5 - 299.1 3.331 - 4.995	24.9 0.416
		•		U/L μkat/L	249.3 4.163	199.5 - 299.1 3.331 - 4.995	24.9 0.416
			•	U/L μkat/L	231.3 3.863	185.1 - 277.5 3.091 - 4.635	23.1 0.386
CA (AZ III) ABX Pentra Calcium AS CP (A11A01954)	•			mmol/L mg/L mg/dL	3.13 125.5 12.55	2.51 - 3.75 100.7 - 150.3 10.07 - 15.03	0.31 12.4 1.24
		•		mmol/L mg/L mg/dL	3.13 125.5 12.55	2.51 - 3.75 100.7 - 150.3 10.07 - 15.03	0.31 12.4 1.24
			•	mmol/L mg/L mg/dL	3.01 120.7 12.07	2.41 - 3.61 96.7 - 144.7 9.67 - 14.47	0.30 12.0 1.20
				mmol/L	3.13	2.51 - 3.75	0.31
				mmol/L	3.01	2.41 - 3.61	0.30

* Confidence range: +/- 2 SD

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD		
CREA (CREA3 / CREA4) (Jaffe/Alkaline Picrate) ABX Pentra Creatinine 120 CP (A11A01933)	•			μmol/L mmol/L mg/dL	13086 13.09 147.9	10468 - 15704 10.47 - 15.71 118.3 - 177.5	1309 1.31 14.8		
		•		μmol/L mmol/L mg/dL	13086 13.09 147.9	10468 - 15704 10.47 - 15.71 118.3 - 177.5	1309 1.31 14.8		
			•	μmol/L mmol/L mg/dL	12474 12.47 141.0	9980 - 14968 9.97 - 14.97 112.8 - 169.2	1247 1.25 14.1		
	CREA ENZ. ABX Pentra Enzymatic Creatinine CP (A11A01907)	•			μmol/L mmol/L mg/dL	14401 14.40 162.7	11521 - 17281 11.52 - 17.28 130.1 - 195.3	1440 1.44 16.3	
			•		μmol/L mmol/L mg/dL	14401 14.40 162.7	11521 - 17281 11.52 - 17.28 130.1 - 195.3	1440 1.44 16.3	
				•	μmol/L mmol/L mg/dL	14147 14.15 159.9	11317 - 16977 11.31 - 16.99 127.9 - 191.9	1415 1.42 16.0	
		GluK (HK) ABX Pentra Glucose HK CP (A11A01667)	•			mmol/L g/L mg/dL	16.07 2.893 289.3	12.85 - 19.29 2.313 - 3.473 231.3 - 347.3	1.61 0.290 29.0
				•		mmol/L g/L mg/dL	16.07 2.893 289.3	12.85 - 19.29 2.313 - 3.473 231.3 - 347.3	1.61 0.290 29.0
					•	mmol/L g/L mg/dL	15.93 2.867 286.7	12.75 - 19.11 2.295 - 3.439 229.5 - 343.9	1.59 0.286 28.6
GluP (PAP) ABX Pentra Glucose PAP CP (A11A01668)	•			mmol/L g/L mg/dL	18.06 3.251 325.1	14.44 - 21.68 2.599 - 3.903 259.9 - 390.3	1.81 0.326 32.6		
		•		mmol/L g/L mg/dL	18.06 3.251 325.1	14.44 - 21.68 2.599 - 3.903 259.9 - 390.3	1.81 0.326 32.6		
			•	mmol/L g/L mg/dL	16.43 2.957 295.7	13.15 - 19.71 2.367 - 3.547 236.7 - 354.7	1.64 0.295 29.5		
	PHOS (Ammonium Molybdate UV) ABX Pentra Phosphorus CP (A11A01665)	•			mmol/L mg/L mg/dL	15.68 486 48.6	12.54 - 18.82 388 - 584 38.8 - 58.4	1.57 49 4.9	
			•		mmol/L mg/L mg/dL	15.68 486 48.6	12.54 - 18.82 388 - 584 38.8 - 58.4	1.57 49 4.9	
				•	mmol/L mg/L mg/dL	15.56 482 48.2	12.44 - 18.68 386 - 578 38.6 - 57.8	1.56 48 4.8	

* Confidence range: +/- 2 SD

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD
TPU (Pyrogallol red) ABX Pentra Urinary Proteins CP (A11A01642)	•			g/L	0.630	0.504 - 0.756	0.063
		•		mg/L	630	504 - 756	63
				mg/dL	63.0	50.4 - 75.6	6.3
				g/L	0.630	0.504 - 0.756	0.063
		•		mg/L	630	504 - 756	63
				mg/dL	63.0	50.4 - 75.6	6.3
UREA (GLDH) ABX Pentra Urea CP (A11A01641)			•	g/L	0.627	0.501 - 0.753	0.063
				mg/L	627	501 - 753	63
				mg/dL	62.7	50.1 - 75.3	6.3
	•			mmol/L	282	226 - 338	28
				mg/L	16920	13560 - 20280	1680
				mg/dL	1692	1356 - 2028	168
BUN (GLDH) ABX Pentra Urea CP (A11A01641)		•		mmol/L	282	226 - 338	28
				mg/L	16920	13560 - 20280	1680
				mg/dL	1692	1356 - 2028	168
			•	mmol/L	258	206 - 310	26
				mg/L	15480	12360 - 18600	1560
				mg/dL	1548	1236 - 1860	156
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)	•			mmol/L	282	226 - 338	28
		•		mg/dL	792	634 - 950	79
				mmol/L	282	226 - 338	28
				mg/dL	792	634 - 950	79
			•	mmol/L	258	206 - 310	26
				mg/dL	725	579 - 871	73
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)	•			μmol/L	814	652 - 976	81
				mmol/L	0.81	0.65 - 0.97	0.08
				mg/dL	13.68	10.96 - 16.40	1.36
		•		μmol/L	814	652 - 976	81
				mmol/L	0.81	0.65 - 0.97	0.08
				mg/dL	13.68	10.96 - 16.40	1.36
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)			•	μmol/L	834	668 - 1000	83
				mmol/L	0.83	0.67 - 0.99	0.08
				mg/dL	14.01	11.23 - 16.79	1.39

* Confidence range: +/- 2 SD