


REF A11A01616

CAL 5 x 1 ml

IVD 



 HORIBA ABX SAS
B.P. 7290
34184 MONTPELLIER Cedex 4
FRANCE

ABX Pentra CRP Cal

- Pentra C 200
- ABX Pentra 400

Calibrator for the measurement of C-Reactive Protein (CRP) by latex-enhanced immunoturbidimetric assay.

Intended Use

ABX Pentra CRP Cal is used to calibrate **ABX Pentra CRP CP**, Ref. A11A01611.

Characteristics

- **ABX Pentra CRP Cal** is a liquid calibrator prepared by diluting C-reactive protein (CRP) with normal human serum at various concentrations.
- **ABX Pentra CRP Cal** is ready-to-use. The kit is composed of 5 vials of 1 ml. Each one has a different concentration (shown on each vial): 2.5, 10, 40, 80 and 160 mg/l. As the color of caps varies according to the CRP levels in the vial, care should be taken not to interchange the caps.
- **ABX Pentra CRP Cal** should be used according to this calibrator notice and as specified in the respective instructions for use of the reagent. The manufacturer cannot guarantee its performance if used otherwise.

Handling

1. Remove the cap of each vial, use a pipette to transfer the required volume into a sample cup.
2. Place the sample cups on the instrument:
 - For **Pentra C 200**: Place each sample cup in the correct position on the instrument sample tray.
 - For **ABX Pentra 400**: Place the sample cups on the appropriate rack of the instrument.

Please, refer to the **ABX Pentra CRP CP** notice for further explanations concerning the use of this calibrator on the instrument.

Materials Required but not Provided

- HORIBA Medical reagents and automated clinical chemistry analyzer.

- Standard laboratory equipment

Assigned Values

The assigned values are based on primary calibration with CRM n°470- CAP/IFCC, lot.91/06-19. Please refer to the vial label for the exact concentration.

Storage and Stability

Calibrators, in unopened vials, are stable up to the expiry date written on the label if stored at 2-10°C and protected from light.

Once opened, **ABX Pentra CRP Cal** is stable for 3 months at 2-10°C.

This stability is obtained when vials are tightly recapped immediately after use and if contamination is avoided. Do not freeze.

Packaging spoiling

In case of protective packaging spoiling, do not use the calibrator if the damages might have an effect on the product performance.

Waste Management

- Please refer to local legal requirements.
- This calibrator contains less than 0.1 % of sodium azide as a preservative. As sodium azide may react with lead and copper to form explosive metal azides, this calibrator should be disposed of by flushing with copious amounts of water.

ABX Pentra CRP Cal

General Precautions

- **ABX Pentra CRP Cal** should be used only for the determination of the calibration curve.
- This calibrator is for professional *in vitro* diagnostic use only.
- Observe the standard laboratory precautions for use.
- **Warning:** Human source material. Treat as potentially infectious. Each plasma donor unit used in the preparation of this product has been tested by an FDA approved method and found nonreactive for the presence of HBsAg, HCV and antibody to HIV1/2. Because no known test method can offer complete assurance that hepatitis B virus, Human Immunodeficiency Virus (HIV) or other infectious agents are absent, the calibrators should be treated like patient specimens as potentially infectious and handled with appropriate cautions in accordance with good laboratory practices (1, 2).
- The calibrator vials should be discarded after use. Disposal of all waste material should be in accordance with local guidelines.
- Please refer to the MSDS associated with the calibrator.
- Do not use the product if there is visible evidence of biological, chemical or physical deterioration.

Warning

It is the user's responsibility to verify that this document is applicable to the calibrator used.

Reference

1. Occupational Safety and Health Standards : bloodborne pathogens. (29 CFR 1910. 1030). Federal Register July 1, 1998 ; **6** :267-280.
2. Council Directive (2000/54/EC). Official Journal of the European Communities. No. L262 from October 17, 2000: 21-45.