

**LOT** MX 440 L  
 Rev 1

**CONTROL** L

(Exp.)

**2023-06-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 BP - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.1			2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.37	2.35			2.37					± 0.17
HB HGB	g/dl	5.9	6.0			5.7					± 0.5
	g/l	59	60			57					± 5
	mmol/l	3.66	3.71			3.52					± 0.31
HT HCT	%	16.6	16.5			17.1					± 2.0
	l/l	0.166	0.165			0.171					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	70	70			72					± 4
TGMH MCH	pg	24.9	25.4			23.9					± 2.0
	fmol	1.55	1.58			1.49					± 0.12
CCMH MCHC	g/dl	35.6	36.3			33.2					± 3.0
	g/l	356	363			332					± 30
	mmol/l	22.08	22.57			20.65					± 1.86
IDR RDW	%	15.5	14.0			15.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	79	84			81					± 22
VPM MPV	µm <sup>3</sup> ; fl	9.0	8.4			8.2					± 2.0
LYMPHO	#	1.36	1.42			1.36					± 0.40
	%	68.0	67.7			68.0					± 8.0
MONO	#	0.13	0.19			0.14					± 0.20
	%	6.5	9.0			7.0					± 6.0
GRANULO	#	0.51	0.49			0.50					± 0.40
	%	25.5	23.5			25.0					± 7.0

<b>LOT</b>		<b>MX 440 N</b>	<b>CONTROL N</b>				(Exp.)		<b>2023-06-05</b> (YYYY-MM-DD)		
Rev 1		<b>Minilysebio</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>7.7</b>	<b>7.6</b>			<b>7.5</b>					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>4.62</b>	<b>4.58</b>			<b>4.59</b>					± 0.26
HB HGB	g/dl	<b>13.7</b>	<b>13.8</b>			<b>13.4</b>					± 0.7
	g/l	<b>137</b>	<b>138</b>			<b>134</b>					± 7
	mmol/l	<b>8.51</b>	<b>8.57</b>			<b>8.34</b>					± 0.43
HT HCT	%	<b>37.4</b>	<b>37.6</b>			<b>38.6</b>					± 2.7
	l/l	<b>0.374</b>	<b>0.376</b>			<b>0.386</b>					± 0.027
VGM MCV	µm <sup>3</sup> ; fl	<b>81</b>	<b>82</b>			<b>84</b>					± 4
TGMH MCH	pg	<b>29.7</b>	<b>30.1</b>			<b>29.3</b>					± 2.0
	fmol	<b>1.84</b>	<b>1.87</b>			<b>1.82</b>					± 0.12
CCMH MCHC	g/dl	<b>36.6</b>	<b>36.7</b>			<b>34.8</b>					± 3.0
	g/l	<b>366</b>	<b>367</b>			<b>348</b>					± 30
	mmol/l	<b>22.73</b>	<b>22.82</b>			<b>21.62</b>					± 1.86
IDR RDW	%	<b>14.5</b>	<b>13.5</b>			<b>14.5</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>257</b>	<b>272</b>			<b>267</b>					± 40
VPM MPV	µm <sup>3</sup> ; fl	<b>8.5</b>	<b>7.9</b>			<b>7.7</b>					± 2.0
LYMPHO	#	<b>2.77</b>	<b>2.82</b>			<b>2.63</b>					± 0.40
	%	<b>36.0</b>	<b>37.1</b>			<b>35.0</b>					± 5.0
MONO	#	<b>0.46</b>	<b>0.47</b>			<b>0.45</b>					± 0.30
	%	<b>6.0</b>	<b>6.2</b>			<b>6.0</b>					± 4.0
GRANULO	#	<b>4.47</b>	<b>4.32</b>			<b>4.43</b>					± 0.60
	%	<b>58.0</b>	<b>56.8</b>			<b>59.0</b>					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143

<b>LOT</b>		<b>MX 440 H</b>	<b>CONTROL H</b>				(Exp.)				<b>2023-06-05</b> (YYYY-MM-DD)
Rev 1		<b>Minilysebio</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.3</b>	<b>20.1</b>			<b>19.8</b>					± 1.7
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.69</b>	<b>5.67</b>			<b>5.76</b>					± 0.32
HB HGB	g/dl	<b>18.5</b>	<b>18.2</b>			<b>18.0</b>					± 0.9
	g/l	<b>185</b>	<b>182</b>			<b>180</b>					± 9
	mmol/l	<b>11.49</b>	<b>11.30</b>			<b>11.19</b>					± 0.56
HT HCT	%	<b>50.6</b>	<b>50.5</b>			<b>52.4</b>					± 3.4
	l/l	<b>0.506</b>	<b>0.505</b>			<b>0.524</b>					± 0.034
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>89</b>			<b>91</b>					± 4
TGMH MCH	pg	<b>32.5</b>	<b>32.1</b>			<b>31.3</b>					± 2.0
	fmol	<b>2.02</b>	<b>1.99</b>			<b>1.94</b>					± 0.12
CCMH MCHC	g/dl	<b>36.5</b>	<b>36.1</b>			<b>34.4</b>					± 3.0
	g/l	<b>365</b>	<b>361</b>			<b>344</b>					± 30
	mmol/l	<b>22.69</b>	<b>22.40</b>			<b>21.35</b>					± 1.86
IDR RDW	%	<b>14.0</b>	<b>14.0</b>			<b>14.5</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>514</b>	<b>537</b>			<b>542</b>					± 65
VPM MPV	µm <sup>3</sup> ; fl	<b>8.5</b>	<b>7.8</b>			<b>7.5</b>					± 2.0
LYMPHO	#	<b>3.65</b>	<b>3.62</b>			<b>3.47</b>					± 1.00
	%	<b>18.0</b>	<b>18.0</b>			<b>17.5</b>					± 5.0
MONO	#	<b>1.02</b>	<b>1.41</b>			<b>0.89</b>					± 0.80
	%	<b>5.0</b>	<b>7.0</b>			<b>4.5</b>					± 4.0
GRANULO	#	<b>15.63</b>	<b>15.07</b>			<b>15.44</b>					± 1.40
	%	<b>77.0</b>	<b>75.0</b>			<b>78.0</b>					± 7.0