


LOT PX 440

CONTROL

 (Exp.) **2023-05-05**
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE				
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE	
		PENTRA						PENTRA						PENTRA						
			DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.9	2.9			± 0.4	8.6	8.6			± 1.0	19.0	19.0			± 2.2				
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.30	2.30			± 0.16	4.56	4.56			± 0.20	5.26	5.26			± 0.25				
HB HGB	g/dl	6.3	6.3			± 0.4	13.8	13.8			± 0.5	16.9	16.9			± 0.6				
	g/l	63	63			± 4	138	138			± 5	169	169			± 6				
	mmol/l	3.91	3.91			± 0.25	8.57	8.57			± 0.31	10.49	10.49			± 0.37				
HT HCT	%	18.9	18.9			± 1.5	40.1	40.1			± 2.0	48.9	48.9			± 2.5				
	l/l	0.189	0.189			± 0.015	0.401	0.401			± 0.020	0.489	0.489			± 0.025				
VGM MCV	µm ³ ; fl	82	82			± 5	88	88			± 5	93	93			± 5				
TGMH MCH	pg	27.4	27.4			± 2.0	30.3	30.3			± 2.0	32.1	32.1			± 2.5				
	fmol	1.70	1.70			± 0.12	1.88	1.88			± 0.12	2.00	2.00			± 0.16				
CCMH MCHC	g/dl	33.4	33.4			± 3.0	34.4	34.4			± 3.0	34.5	34.5			± 3.0				
	g/l	334	334			± 30	344	344			± 30	345	345			± 30				
	mmol/l	20.74	20.74			± 1.86	21.36	21.36			± 1.86	21.45	21.45			± 1.86				
IDR RDW	%	16.5	16.5			± 4.0	14.5	14.5			± 4.0	14.0	14.0			± 4.0				
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	78	78			± 20	266	266			± 30	498	498			± 50				
VPM MPV	µm ³ ; fl	8.9	8.9			± 2.0	8.6	8.6			± 2.0	9.5	9.5			± 2.0				
NEUT	#	1.56	1.56			± 0.35	4.86	4.86			± 0.90	13.70	13.70			± 1.90				
	%	53.7	53.7			± 10.0	56.5	56.5			± 10.0	72.3	72.3			± 10.0				
LYMPHO	#	0.86	0.86			± 0.33	2.80	2.80			± 0.70	3.10	3.10			± 1.50				
	%	29.5	29.5			± 12.0	32.6	32.6			± 8.0	16.3	16.3			± 8.0				
MONO	#	0.12	0.12			± 0.12	0.20	0.20			± 0.20	0.42	0.42			± 0.42				
	%	4.3	4.3			± 4.3	2.3	2.3			± 2.3	2.2	2.2			± 2.2				
EOS	#	0.22	0.22			± 0.22	0.34	0.34			± 0.34	0.84	0.84			± 0.84				
	%	7.6	7.6			± 7.6	4.0	4.0			± 4.0	4.4	4.4			± 4.4				
BASO	#	0.14	0.14			± 0.14	0.40	0.40			± 0.40	0.91	0.91			± 0.91				
	%	4.9	4.9			± 4.9	4.6	4.6			± 4.6	4.8	4.8			± 4.8				

1300090306
BACK / VERSO
Ref: TEMP-0821 Rev.47