

LOT		MX 440 L		CONTROL		L		 (Exp.)		2023-06-05 (YYYY - MM - DD)											
PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse								TOLERANCES TOLERANCE									
				MICROS		MICROS		MICROS		MICROS				ARGOS - HELIOS		PENTRA		PENTRA		PENTRA	
				ES60 Care ST		CRP 200				60 8P - 60 18P ADVIA 60											
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1			2.0											± 0.4			
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.36	2.35			2.36											± 0.17			
HB	HGB	g/dl	6.1	6.1			6.1											± 0.5			
		g/l	61	61			61										± 5				
		mmol/l	3.79	3.79			3.79										± 0.31				
HT	HCT	%	16.5	16.5			16.5										± 2.0				
		l/l	0.165	0.165			0.165										± 0.020				
VGM	MCV	µm ³ ; fl	70	70			70											± 4			
TGMH	MCH	pg	25.8	26.0			25.8											± 2.0			
		fmol	1.61	1.61			1.61											± 0.12			
CCMH	MCHC	g/dl	36.9	37.1			36.9											± 3.0			
		g/l	369	371			369											± 30			
		mmol/l	22.93	23.03			22.93											± 1.86			
IDR	RDW	%	15.0	14.0			14.0											± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	78	84			81											± 22			
VPM	MPV	µm ³ ; fl	8.8	8.4			8.2											± 2.0			
LYMPHO	#	1.37	1.41			1.32												± 0.40			
	%	65.0	67.0			66.0												± 8.0			
MONO	#	0.20	0.19			0.18												± 0.20			
	%	9.5	9.0			9.0												± 6.0			
GRANULO	#	0.54	0.50			0.50												± 0.40			
	%	25.5	24.0			25.0												± 7.0			

LOT		MX 440 N		Rev 1		CONTROL N		(Exp.)		2023-06-05 (YYYY - MM - DD)	
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.6			7.7					± 0.8
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	4.57	4.58			4.61					± 0.26
HB HGB	g/dl	13.6	13.8			13.7					± 0.7
	g/l	136	138			137					± 7
	mmol/l	8.45	8.57			8.51					± 0.43
HT HCT	%	37.5	37.6			37.8					± 2.7
	l/l	0.375	0.376			0.378					± 0.027
VGM MCV	µm ³ ; fl	82	82			82					± 4
TGMH MCH	pg	29.8	30.1			29.7					± 2.0
	fmol	1.85	1.87			1.85					± 0.12
CCMH MCHC	g/dl	36.3	36.7			36.2					± 3.0
	g/l	363	367			362					± 30
	mmol/l	22.54	22.82			22.51					± 1.86
IDR RDW	%	14.5	13.5			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	272	272			267					± 40
VPM MPV	µm ³ ; fl	8.5	7.9			7.7					± 2.0
LYMPHO	#	2.73	2.66			2.66					± 0.40
	%	35.5	35.0			34.5					± 5.0
MONO	#	0.50	0.53			0.50					± 0.30
	%	6.5	7.0			6.5					± 4.0
GRANULO	#	4.47	4.41			4.54					± 0.60
	%	58.0	58.0			59.0					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104144

LOT		MX 440 H		Rev 1		CONTROL H		(Exp.)		2023-06-05 (YYYY - MM - DD)	
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.5	20.1			20.4					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.71	5.67			5.74					± 0.32
HB HGB	g/dl	18.2	18.2			18.2					± 0.9
	g/l	182	182			182					± 9
	mmol/l	11.30	11.30			11.30					± 0.56
HT HCT	%	50.8	50.5			51.7					± 3.4
	l/l	0.508	0.505			0.517					± 0.034
VGM MCV	µm ³ ; fl	89	89			90					± 4
TGMH MCH	pg	31.9	32.1			31.7					± 2.0
	fmol	1.98	1.99			1.97					± 0.12
CCMH MCHC	g/dl	35.8	36.1			35.2					± 3.0
	g/l	358	361			352					± 30
	mmol/l	22.24	22.40			21.88					± 1.86
IDR RDW	%	14.5	14.0			13.0					± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	553	537			542					± 65
VPM MPV	µm ³ ; fl	8.2	7.8			7.5					± 2.0
LYMPHO	#	3.28	3.42			3.26					± 1.00
	%	16.0	17.0			16.0					± 5.0
MONO	#	1.23	1.31			1.02					± 0.80
	%	6.0	6.5			5.0					± 4.0
GRANULO	#	15.99	15.38			16.12					± 1.40
	%	78.0	76.5			79.0					± 7.0