

<b>LOT</b>		<b>MX 440 L</b>		<b>CONTROL L</b>								(Exp.) <b>2023-06-05</b> (YYYY - MM - DD)				
		Rev 1		<b>ABX Lyse</b>												
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS		MICROS		MICROS		MICROS		ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP 200					60 8P - 60 18P ADVIA 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1	2.1					2.0								± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.36	2.35					2.36								± 0.17
HB HGB	g/dl	6.1	6.1					6.1								± 0.5
	g/l	61	61					61								± 5
	mmol/l	3.79	3.79					3.79								± 0.31
HT HCT	%	16.5	16.5					16.5								± 2.0
	l/l	0.165	0.165					0.165								± 0.020
VGM MCV	µm <sup>3</sup> ; fl	70	70					70								± 4
TGMH MCH	pg	25.8	26.0					25.8								± 2.0
	fmol	1.61	1.61					1.61								± 0.12
CCMH MCHC	g/dl	36.9	37.1					36.9								± 3.0
	g/l	369	371					369								± 30
	mmol/l	22.93	23.03					22.93								± 1.86
IDR RDW	%	15.0	14.0					14.0								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	78	84					81								± 22
VPM MPV	µm <sup>3</sup> ; fl	8.8	8.4					8.2								± 2.0
LYMPHO	#	1.37	1.41					1.32								± 0.40
	%	65.0	67.0					66.0								± 8.0
MONO	#	0.20	0.19					0.18								± 0.20
	%	9.5	9.0					9.0								± 6.0
GRANULO	#	0.54	0.50					0.50								± 0.40
	%	25.5	24.0					25.0								± 7.0

LOT		MX 440 N		CONTROL N						(Exp.) <b>2023-06-05</b> (YYYY-MM-DD)	
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.7	7.6			7.7					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.57	4.58			4.61					± 0.26
HB HGB	g/dl	13.6	13.8			13.7					± 0.7
	g/l	136	138			137					± 7
	mmol/l	8.45	8.57			8.51					± 0.43
HT HCT	%	37.5	37.6			37.8					± 2.7
	l/l	0.375	0.376			0.378					± 0.027
VGM MCV	µm <sup>3</sup> ; fl	82	82			82					± 4
TGMH MCH	pg	29.8	30.1			29.7					± 2.0
	fmol	1.85	1.87			1.85					± 0.12
CCMH MCHC	g/dl	36.3	36.7			36.2					± 3.0
	g/l	363	367			362					± 30
	mmol/l	22.54	22.82			22.51					± 1.86
IDR RDW	%	14.5	13.5			13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	272	272			267					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.5	7.9			7.7					± 2.0
LYMPHO	#	2.73	2.66			2.66					± 0.40
	%	35.5	35.0			34.5					± 5.0
MONO	#	0.50	0.53			0.50					± 0.30
	%	6.5	7.0			6.5					± 4.0
GRANULO	#	4.47	4.41			4.54					± 0.60
	%	58.0	58.0			59.0					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104144

**LOT**

**MX 440 H**  
Rev 1

**CONTROL H**

(Exp.)

**2023-06-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.5</b>	<b>20.1</b>			<b>20.4</b>					± 1.7
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.71</b>	<b>5.67</b>			<b>5.74</b>					± 0.32
HB HGB	g/dl	<b>18.2</b>	<b>18.2</b>			<b>18.2</b>					± 0.9
	g/l	<b>182</b>	<b>182</b>			<b>182</b>					± 9
	mmol/l	<b>11.30</b>	<b>11.30</b>			<b>11.30</b>					± 0.56
HT HCT	%	<b>50.8</b>	<b>50.5</b>			<b>51.7</b>					± 3.4
	l/l	<b>0.508</b>	<b>0.505</b>			<b>0.517</b>					± 0.034
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>89</b>			<b>90</b>					± 4
TGMH MCH	pg	<b>31.9</b>	<b>32.1</b>			<b>31.7</b>					± 2.0
	fmol	<b>1.98</b>	<b>1.99</b>			<b>1.97</b>					± 0.12
CCMH MCHC	g/dl	<b>35.8</b>	<b>36.1</b>			<b>35.2</b>					± 3.0
	g/l	<b>358</b>	<b>361</b>			<b>352</b>					± 30
	mmol/l	<b>22.24</b>	<b>22.40</b>			<b>21.88</b>					± 1.86
IDR RDW	%	<b>14.5</b>	<b>14.0</b>			<b>13.0</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>553</b>	<b>537</b>			<b>542</b>					± 65
VPM MPV	µm <sup>3</sup> ; fl	<b>8.2</b>	<b>7.8</b>			<b>7.5</b>					± 2.0
LYMPHO	#	<b>3.28</b>	<b>3.42</b>			<b>3.26</b>					± 1.00
	%	<b>16.0</b>	<b>17.0</b>			<b>16.0</b>					± 5.0
MONO	#	<b>1.23</b>	<b>1.31</b>			<b>1.02</b>					± 0.80
	%	<b>6.0</b>	<b>6.5</b>			<b>5.0</b>					± 4.0
GRANULO	#	<b>15.99</b>	<b>15.38</b>			<b>16.12</b>					± 1.40
	%	<b>78.0</b>	<b>76.5</b>			<b>79.0</b>					± 7.0