



LOT

MX 440 L

CONTROL

L

(Exp.)

2023-06-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			2.1						± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			2.38						± 0.15
HB HGB	g/dl			6.1						± 0.4
	g/l			61						± 4
	mmol/l			3.79						± 0.25
HT HCT	%			16.4						± 2.0
	l/l			0.164						± 0.020
VGM MCV	µm ³ ; fl			69						± 4
TGMH MCH	pg			25.6						± 2.0
	fmol			1.59						± 0.12
CCMH MCHC	g/dl			37.1						± 3.0
	g/l			371						± 30
	mmol/l			23.07						± 1.86
IDR RDW	%			14.5						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			73						± 20
VPM MPV	µm ³ ; fl			8.0						± 2.0
LYMPHO	#			1.37						± 0.40
	%			65.0						± 8.0
MONO	#			0.17						± 0.20
	%			8.0						± 6.0
GRANULO	#			0.57						± 0.40
	%			27.0						± 7.0

LOT

MX 440 N
Rev 1

CONTROL

N

 (Exp.)

2023-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			7.9						± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			4.55						± 0.18
HB HGB	g/dl			13.5						± 0.5
	g/l			135						± 5
	mmol/l			8.38						± 0.31
HT HCT	%			36.9						± 2.5
	l/l			0.369						± 0.025
VGM MCV	µm ³ ; fl			81						± 4
TGMH MCH	pg			29.7						± 2.0
	fmol			1.84						± 0.12
CCMH MCHC	g/dl			36.6						± 3.0
	g/l			366						± 30
	mmol/l			22.75						± 1.86
IDR RDW	%			13.0						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			272						± 40
VPM MPV	µm ³ ; fl			8.3						± 2.0
LYMPHO	#			2.80						± 0.40
	%			35.5						± 5.0
MONO	#			0.55						± 0.30
	%			7.0						± 4.0
GRANULO	#			4.54						± 0.60
	%			57.5						± 6.0

Ret: TEMP-0830 Rev.42 BACK / VERSO 1300104144

LOT

MX 440 H
Rev 1

CONTROL

H

 (Exp.)

2023-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			20.5						± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			5.59						± 0.20
HB HGB	g/dl			17.8						± 0.6
	g/l			178						± 6
	mmol/l			11.05						± 0.37
HT HCT	%			49.2						± 3.0
	l/l			0.492						± 0.030
VGM MCV	µm ³ ; fl			88						± 4
TGMH MCH	pg			31.8						± 2.0
	fmol			1.98						± 0.12
CCMH MCHC	g/dl			36.2						± 3.0
	g/l			362						± 30
	mmol/l			22.47						± 1.86
IDR RDW	%			12.5						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			538						± 45
VPM MPV	µm ³ ; fl			8.3						± 2.0
LYMPHO	#			3.69						± 1.00
	%			18.0						± 5.0
MONO	#			1.13						± 0.80
	%			5.5						± 4.0
GRANULO	#			15.68						± 1.40
	%			76.5						± 7.0