

ABX Erytrol



LOT

NX 440

Rev 1

CONTROL



(Exp.) 2023-05-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA	DX120		DX NEXUS	DX120	DX NEXUS		DX120	DX NEXUS		
		DX120	DX NEXUS										
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	7.3	7.3		± 0.8	17.5	17.5		± 1.5	2.4	2.4		± 0.4
GBC CWBC	10 ³ /mm ³ ; 10 ⁹ /l	7.0	7.0		± 0.8	16.2	16.2		± 1.4	1.4	1.4		± 0.4
ERB	%	3.5	3.5		± 1.7	7.9	7.9		± 3.3	75.5	75.5		± 25.0
ERB	#	0.245	0.245		± 0.100	1.280	1.280		± 0.700	1.057	1.057		± 0.300

1219930076-C

Ref: TEMP-0822 Rev.35

GBC / CWBC

* les valeurs sont données à titre d'indication et doivent être utilisées comme RuO (Utilisation en Recherche Seulement) Indicative expected values, must be used as RuO (Research used only)