

# ABX Minotrol Retic



**LOT** RX 440  
Rev 1

**CONTROL**

(Exp.) 2023-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.3	1.3	± 0.7	5.5	5.5	± 2.2	11.9	11.9	± 3.6			
RETIC	%	1.46	1.46	± 0.73	5.00	5.00	± 1.25	9.70	9.70	± 1.94			
RETIC	10 <sup>6</sup> /µl	0.050	0.050	± 0.025	0.169	0.169	± 0.042	0.325	0.325	± 0.065			
MFI MRI *	%	26.3	26.3	± 8.0	44.0	44.0	± 10.0	50.4	50.4	± 12.0			
GR RBC	10 <sup>6</sup> /µl	3.40	3.40	± 0.30	3.38	3.38	± 0.30	3.35	3.35	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.42 FRONT / RECTO 1300090305

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 440  
Rev 1

**CONTROL**

(Exp.) 2023-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	<b>1.3</b>			± 0.7	<b>5.5</b>			± 2.2	<b>11.9</b>			± 3.6
RETIC	%	<b>1.71</b>			± 0.86	<b>5.11</b>			± 1.28	<b>10.36</b>			± 2.07
RETIC	10 <sup>6</sup> /µl	<b>0.059</b>			± 0.030	<b>0.172</b>			± 0.043	<b>0.348</b>			± 0.070
MFI MRI *	%	<b>29.8</b>			± 8.0	<b>39.0</b>			± 10.0	<b>41.4</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.43</b>			± 0.30	<b>3.37</b>			± 0.30	<b>3.36</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.42 BACK / VERSO 1300090305

\* For research use only in Canada and USA