


LOT PX 440
Rev 2


CONTROL

 (Exp.) 2023-05-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0			± 0.4	8.5	8.5			± 1.0	18.6	18.6			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.26	2.26			± 0.16	4.46	4.46			± 0.20	5.15	5.15			± 0.25			
HB HGB	g/dl	6.2	6.2			± 0.4	13.7	13.7			± 0.5	17.0	17.0			± 0.6			
	g/l	62	62			± 4	137	137			± 5	170	170			± 6			
	mmol/l	3.85	3.85			± 0.25	8.51	8.51			± 0.31	10.56	10.56			± 0.37			
HT HCT	%	18.5	18.5			± 1.5	39.2	39.2			± 2.0	47.9	47.9			± 2.5			
	l/l	0.185	0.185			± 0.015	0.392	0.392			± 0.020	0.479	0.479			± 0.025			
VGM MCV	µm ³ ; fl	82	82			± 5	88	88			± 5	93	93			± 5			
TGMH MCH	pg	27.4	27.4			± 2.0	30.7	30.7			± 2.0	33.0	33.0			± 2.5			
	fmol	1.70	1.70			± 0.12	1.91	1.91			± 0.12	2.05	2.05			± 0.16			
CCMH MCHC	g/dl	33.5	33.5			± 3.0	34.9	34.9			± 3.0	35.5	35.5			± 3.0			
	g/l	335	335			± 30	349	349			± 30	355	355			± 30			
	mmol/l	20.78	20.78			± 1.86	21.68	21.68			± 1.86	22.04	22.04			± 1.86			
IDR RDW	%	13.5	13.5			± 4.0	13.0	13.0			± 4.0	12.5	12.5			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	77	77			± 20	265	265			± 30	488	488			± 50			
VPM MPV	µm ³ ; fl	8.8	8.8			± 2.0	8.8	8.8			± 2.0	9.8	9.8			± 2.0			
NEUT	#	1.45	1.45			± 0.35	4.41	4.41			± 0.90	13.08	13.08			± 1.90			
	%	48.4	48.4			± 10.0	51.9	51.9			± 10.0	70.3	70.3			± 10.0			
LYMPHO	#	1.17	1.17			± 0.33	3.35	3.35			± 0.70	3.53	3.53			± 1.50			
	%	39.0	39.0			± 12.0	39.4	39.4			± 8.0	19.0	19.0			± 8.0			
MONO	#	0.10	0.10			± 0.10	0.20	0.20			± 0.20	0.37	0.37			± 0.37			
	%	3.4	3.4			± 3.4	2.3	2.3			± 2.3	2.0	2.0			± 2.0			
EOS	#	0.20	0.20			± 0.20	0.31	0.31			± 0.31	0.87	0.87			± 0.87			
	%	6.5	6.5			± 6.5	3.6	3.6			± 3.6	4.7	4.7			± 4.7			
BASO	#	0.08	0.08			± 0.08	0.24	0.24			± 0.24	0.74	0.74			± 0.74			
	%	2.7	2.7			± 2.7	2.8	2.8			± 2.8	4.0	4.0			± 4.0			

LOT PX 440
Rev 2

CONTROL

 (Exp.) 2023-05-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.9	2.9	2.9		± 0.4	8.5	8.5	8.4		± 1.0	19.2	18.8	18.6		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.32	2.30	2.30		± 0.16	4.51	4.49	4.47		± 0.20	5.16	5.07	5.10		± 0.25			
HB HGB	g/dl	6.2	6.2	6.2		± 0.4	13.8	13.7	13.7		± 0.5	17.0	16.8	16.9		± 0.6			
	g/l	62	62	62		± 4	138	137	137		± 5	170	168	169		± 6			
HT HCT	mmol/l	3.85	3.85	3.85		± 0.25	8.57	8.51	8.51		± 0.31	10.56	10.43	10.49		± 0.37			
	%	18.3	17.9	18.1		± 1.5	39.2	39.1	39.1		± 2.0	48.5	47.7	47.7		± 2.5			
VGM MCV	l/l	0.183	0.179	0.181		± 0.015	0.392	0.391	0.391		± 0.020	0.485	0.477	0.477		± 0.025			
	µm ³ ·fl	79	78	78.5		± 5	87	87	87.5		± 5	94	94	93.5		± 5			
TGMH MCH	pg	26.7	27.0	27.0		± 2.0	30.6	30.5	30.6		± 2.0	32.9	33.1	33.1		± 2.5			
	fmol	1.66	1.67	1.67		± 0.12	1.90	1.89	1.90		± 0.12	2.05	2.06	2.06		± 0.16			
CCMH MCHC	g/dl	33.8	34.6	34.3		± 3.0	35.2	35.1	35.0		± 3.0	35.0	35.3	35.4		± 3.0			
	g/l	338	346	343		± 30	352	351	350		± 30	350	353	354		± 30			
IDR RDW	mmol/l	21.01	21.46	21.32		± 1.86	21.84	21.78	21.75		± 1.86	21.77	21.89	22.01		± 1.86			
	%	13.0	13.5	13.5		± 4.0	13.0	12.5	11.5		± 4.0	12.0	12.0	11.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	80	75	77		± 20	265	263	254		± 30	493	496	468		± 50			
VPM MPV	µm ³ ·fl	8.4	8.5	8.0		± 2.0	8.1	8.3	8.0		± 2.0	8.9	9.1	8.5		± 2.0			
NEUT	#	1.39	1.38	1.44		± 0.35	4.33	4.34	4.41		± 0.90	13.42	13.22	13.11		± 1.90			
	%	47.8	47.6	49.5		± 10.0	50.9	51.0	52.5		± 10.0	69.9	70.3	70.5		± 10.0			
LYMPHO	#	1.16	1.16	1.12		± 0.33	3.43	3.44	3.32		± 0.70	3.74	3.68	3.53		± 1.50			
	%	40.0	39.9	38.5		± 12.0	40.3	40.5	39.5		± 8.0	19.5	19.6	19.0		± 8.0			
MONO	#	0.11	0.10	0.10		± 0.10	0.20	0.18	0.17		± 0.17	0.40	0.34	0.28		± 0.28			
	%	3.7	3.5	3.5		± 3.5	2.4	2.1	2.0		± 2.0	2.1	1.8	1.5		± 1.5			
EOS	#	0.17	0.19	0.16		± 0.16	0.31	0.31	0.25		± 0.25	0.84	0.81	0.93		± 0.81			
	%	5.8	6.4	5.5		± 5.5	3.6	3.6	3.0		± 3.0	4.4	4.3	5.0		± 4.3			
BASO	#	0.08	0.08	0.09		± 0.08	0.24	0.24	0.25		± 0.24	0.79	0.75	0.74		± 0.74			
	%	2.7	2.6	3.0		± 2.6	2.8	2.8	3.0		± 2.8	4.1	4.0	4.0		± 4.0			

1300090306 FRONT / RECTO Ref: TEMP-0821 Rev.47