

# ABX Minotrol CRP



**LOT**

**MC 427**

**CONTROL**



(Exp.) **2023-07-05**  
(YYYY-MM-DD)

Rev 1

		ABX Lyse														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	2.0	2.0			± 0.4	7.9	7.9			± 1.0	20.0	20.0			± 2.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	2.34	2.32			± 0.15	4.47	4.45			± 0.18	5.67	5.61			± 0.25
HB HGB	g/dl	6.1	6.0			± 0.4	13.4	13.4			± 0.5	18.3	18.3			± 0.7
	g/l	61	60			± 4	134	134			± 5	183	183			± 7
HT HCT	mmol/l	3.79	3.73			± 0.25	8.32	8.32			± 0.31	11.36	11.36			± 0.44
	%	16.6	16.2			± 2.0	36.7	36.5			± 2.5	51.6	51.1			± 3.0
VGM MCV	l/l	0.166	0.162			± 0.020	0.367	0.365			± 0.025	0.516	0.511			± 0.030
	µm <sup>3</sup> ; fl	71	70			± 4	82	82			± 4	91	91			± 4
TGMH MCH	pg	26.1	25.9			± 2.4	30.0	30.1			± 2.4	32.3	32.6			± 2.8
	fmol	1.62	1.61			± 0.15	1.86	1.87			± 0.15	2.00	2.03			± 0.18
CCMH MCHC	g/dl	36.7	36.9			± 3.0	36.6	36.7			± 3.0	35.5	35.8			± 3.0
	g/l	367	369			± 30	366	367			± 30	355	358			± 30
IDR RDW	mmol/l	22.80	22.94			± 1.86	22.70	22.80			± 1.86	22.03	22.26			± 1.86
	%	14.0	14.0			± 3.0	13.5	13.0			± 3.0	13.0	12.5			± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	76	72			± 20	256	254			± 40	509	515			± 60
VPM MPV	µm <sup>3</sup> ; fl	8.7	7.8			± 2.0	8.2	7.4			± 2.0	7.9	7.3			± 2.0
LYMPHO	#	1.4	1.3			± 0.2	2.6	2.6			± 0.5	3.4	3.3			± 1.1
	%	68.0	66.0			± 8.0	33.5	32.5			± 6.0	17.0	16.5			± 5.0
MONO	#	0.2	0.2			± 0.2	0.6	0.6			± 0.4	1.1	0.9			± 1.0
	%	9.5	9.0			± 7.0	8.0	7.0			± 5.0	5.5	4.5			± 5.0
GRANULO	#	0.5	0.5			± 0.2	4.6	4.8			± 0.5	15.5	15.8			± 1.5
	%	22.5	25.0			± 7.0	58.5	60.5			± 6.0	77.5	79.0			± 7.0
CRP	mg/dl	0.30	N/A			± 0.30	2.08	N/A			± 0.70	4.08	N/A			± 1.00
	mg/l	3.0	N/A			± 3.0	20.8	N/A			± 7.0	40.8	N/A			± 10.0

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170