

# ABX Minocal

**LOT**

**CX 481**

**CAL**



**(Exp.) 2023-05-05**

Rev **2**

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P 60 18P					
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>8.8</b>	<b>8.8</b>			<b>8.9</b>					± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.39</b>	<b>4.38</b>			<b>4.37</b>					± 0.06
HB HGB	g/dl	<b>13.0</b>	<b>12.6</b>			<b>12.9</b>					± 0.2
	g/l	<b>130</b>	<b>126</b>			<b>129</b>					± 2
	mmol/l	<b>8.07</b>	<b>7.85</b>			<b>8.01</b>					± 0.12
HT HCT	%	<b>36.4</b>	<b>36.4</b>			<b>36.9</b>					± 1.0
	l/l	<b>0.364</b>	<b>0.364</b>			<b>0.369</b>					± 0.010
PLA PLT	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>225</b>	<b>244</b>			<b>246</b>					± 10
VMP MPV	µm <sup>3</sup> ; fl	<b>9.3</b>	<b>8.6</b>			<b>9.2</b>					± 0.5

# ABX Minocal

**LOT**

**CX 481**

**CAL**



**(Exp.) 2023-05-05**  
(YYYY-MM-DD)

Rev **2**

PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>8.8</b>	<b>8.6</b>							± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.40</b>	<b>4.40</b>							± 0.06
HB HGB	g/dl	<b>12.7</b>	<b>12.7</b>							± 0.2
	g/l	<b>127</b>	<b>127</b>							± 2
	mmol/l	<b>7.89</b>	<b>7.89</b>							± 0.12
HT HCT	%	<b>36.3</b>	<b>36.5</b>							± 1.0
	l/l	<b>0.363</b>	<b>0.365</b>							± 0.010
PLA PLT	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>222</b>	<b>237</b>							± 10
VMP MPV	µm <sup>3</sup> ; fl	N/A	<b>8.9</b>							± 0.5

Ref: TEMP-0387 Rev.48 FRONT / RECTO 1300105208