

ABX Minotrol CRP



LOT

MC 427

CONTROL



(Exp.) 2023-07-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														TOLERANCES TOLERANCE	
				CONTROL		1		CONTROL		2		CONTROL		3					
				MICROS		PENTRA		MICROS		PENTRA		MICROS		PENTRA					
				EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP					
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0	2.0		± 0.4	8.0	8.1	8.5		± 1.0	20.3	20.6	21.3		± 2.2		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.37	2.33	2.35		± 0.15	4.46	4.44	4.50		± 0.18	5.48	5.51	5.61		± 0.25		
HB	HGB	g/dl	6.1	6.1	5.9		± 0.4	13.3	13.4	13.4		± 0.5	18.0	18.1	18.5		± 0.7		
		g/l	61	61	59		± 4	133	134	134		± 5	180	181	185		± 7		
		mmol/l	3.79	3.79	3.66		± 0.25	8.26	8.32	8.32		± 0.31	11.18	11.24	11.49		± 0.44		
HT	HCT	%	16.6	16.5	16.2		± 2.0	36.1	36.4	36.5		± 2.5	49.9	49.6	49.9		± 3.0		
		l/l	0.166	0.165	0.162		± 0.020	0.361	0.364	0.365		± 0.025	0.499	0.496	0.499		± 0.030		
VGM	MCV	µm ³ ; fl	70	71	69.0		± 4	81	82	81.0		± 4	91	90	89.0		± 4		
TGMH	MCH	pg	25.7	26.2	25.1		± 2.4	29.8	30.2	29.8		± 2.4	32.8	32.8	33.0		± 2.8		
		fmol	1.60	1.63	1.56		± 0.15	1.85	1.87	1.85		± 0.15	2.04	2.04	2.05		± 0.18		
CCMH	MCHC	g/dl	36.8	36.9	36.4		± 3.0	36.8	36.8	36.8		± 3.0	36.1	36.5	37.1		± 3.0		
		g/l	368	369	364		± 30	368	368	368		± 30	361	365	371		± 30		
		mmol/l	22.83	22.90	22.60		± 1.86	22.86	22.86	22.83		± 1.86	22.42	22.67	23.01		± 1.86		
IDR	RDW	%	14.0	13.5	14.5		± 3.0	13.0	12.5	13.0		± 3.0	12.5	12.0	12.0		± 3.0		
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	68	67	58		± 20	253	244	247		± 40	503	490	521		± 60		
VPM	MPV	µm ³ ; fl	7.8	8.7	9.5		± 2.0	8.1	8.4	9.2		± 2.0	8.0	8.2	8.8		± 2.0		
LYMPHO		#	1.3	1.3	N/A		± 0.2	2.6	2.6	N/A		± 0.5	3.6	3.4	N/A		± 1.1		
		%	65.5	65.5	N/A		± 8.0	33.0	32.0	N/A		± 6.0	17.5	16.5	N/A		± 5.0		
MONO		#	0.2	0.2	N/A		± 0.2	0.6	0.6	N/A		± 0.4	1.1	1.0	N/A		± 1.0		
		%	8.0	8.5	N/A		± 7.0	7.0	7.5	N/A		± 5.0	5.5	5.0	N/A		± 5.0		
GRANULO		#	0.5	0.5	N/A		± 0.2	4.8	4.9	N/A		± 0.5	15.6	16.2	N/A		± 1.5		
		%	26.5	26.0	N/A		± 7.0	60.0	60.5	N/A		± 6.0	77.0	78.5	N/A		± 7.0		
CRP		mg/dl	0.41	0.48	0.44		± 0.40	2.77	2.80	2.59		± 0.80	5.15	5.21	4.95		± 1.50		
		mg/l	4.1	4.8	4.4		± 4.0	27.7	28.0	25.9		± 8.0	51.5	52.1	49.5		± 15.0		

Ref: TEMP-1070 Rev.42 BACK/VERSO 1300092170