

Yumizen G PT 5 for Yumizen G Line



Reference No.:		1300036338		
Lot No.:		930106 (rev. 1)		
Expiry date:		2025-01-31		
Coagulometers		ISI	MASTER CURVE	
Optical	Yumizen G Line		%	sec
		100	12,7	
		50	17,0	
		33	21,5	
		25	25,7	
		1,18	MNPT	12,7
Important notice				
<ul style="list-style-type: none"> - The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT. - The efficiency of the reconstitution of the reagent may be improved by performing it at 37°C. 				