

# Yumizen G PT Liq 4 for Yumizen G Line



<b>Reference No.:</b>		<b>1300036373</b>		
<b>Lot No.:</b>		<b>920817 (rev. 1)</b>		
<b>Expiry date:</b>		<b>2024-02-29</b>		
<b>Coagulometers</b>		<b>ISI</b>	<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G Line</b>		<b>1,19</b>	<b>%</b>
			<b>100</b>	<b>12,4</b>
			<b>50</b>	<b>17,0</b>
			<b>33</b>	<b>21,6</b>
			<b>25</b>	<b>26,0</b>
		<b>MNPT</b>	<b>12,4</b>	
<b>Important notice</b>				
The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.				