

**LOT** MX 441 L  
 Rev 1

**CONTROL** L

(Exp.)

**2023-08-05**  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.0			2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.29	2.32			2.29					± 0.17
HB HGB	g/dl	6.0	6.0			5.7					± 0.5
	g/l	60	60			57					± 5
	mmol/l	3.73	3.71			3.52					± 0.31
HT HCT	%	16.5	16.2			16.5					± 2.0
	l/l	0.165	0.162			0.165					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	72	70			72					± 4
TGMH MCH	pg	26.2	25.8			24.8					± 2.0
	fmol	1.63	1.60			1.54					± 0.12
CCMH MCHC	g/dl	36.4	36.8			34.4					± 3.0
	g/l	364	368			344					± 30
	mmol/l	22.60	22.86			21.37					± 1.86
IDR RDW	%	15.0	13.5			15.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	85			80					± 22
VPM MPV	µm <sup>3</sup> ; fl	8.6	8.0			7.8					± 2.0
LYMPHO	#	1.34	1.34			1.33					± 0.40
	%	67.0	67.2			66.5					± 8.0
MONO	#	0.16	0.19			0.17					± 0.20
	%	8.0	9.5			8.5					± 6.0
GRANULO	#	0.50	0.47			0.50					± 0.40
	%	25.0	23.5			25.0					± 7.0

**LOT** MX 441 N  
 Rev 1

**CONTROL** N

(Exp.)

**2023-08-05**  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.5	7.4			7.3					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.50	4.52			4.49					± 0.26
HB HGB	g/dl	13.4	13.4			13.2					± 0.7
	g/l	134	134			132					± 7
	mmol/l	8.32	8.32			8.22					± 0.43
HT HCT	%	36.9	36.6			37.3					± 2.7
	l/l	0.369	0.366			0.373					± 0.027
VGM MCV	µm <sup>3</sup> ; fl	82	81			83					± 4
TGMH MCH	pg	29.8	29.6			29.5					± 2.0
	fmol	1.85	1.84			1.83					± 0.12
CCMH MCHC	g/dl	36.3	36.6			35.5					± 3.0
	g/l	363	366			355					± 30
	mmol/l	22.55	22.73			22.05					± 1.86
IDR RDW	%	14.5	13.5			14.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	244	254			260					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.5	8.0			7.4					± 2.0
LYMPHO	#	2.70	2.75			2.59					± 0.40
	%	36.0	37.1			35.5					± 5.0
MONO	#	0.49	0.53			0.51					± 0.30
	%	6.5	7.1			7.0					± 4.0
GRANULO	#	4.31	4.13			4.20					± 0.60
	%	57.5	55.9			57.5					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143

**LOT** MX 441 H  
Rev 1

**CONTROL** H

 (Exp.)

**2023-08-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	19.3	19.1			19.0					± 1.7
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.51	5.60			5.56					± 0.32
HB HGB	g/dl	18.1	17.8			17.7					± 0.9
	g/l	181	178			177					± 9
	mmol/l	11.24	11.05			11.00					± 0.56
HT HCT	%	49.6	49.8			50.6					± 3.4
	l/l	0.496	0.498			0.506					± 0.034
VGM MCV	µm <sup>3</sup> ; fl	90	89			91					± 4
TGMH MCH	pg	32.8	31.8			31.9					± 2.0
	fmol	2.04	1.97			1.98					± 0.12
CCMH MCHC	g/dl	36.5	35.7			35.0					± 3.0
	g/l	365	357			350					± 30
	mmol/l	22.67	22.18			21.75					± 1.86
IDR RDW	%	14.0	13.5			13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	492	507			520					± 65
VPM MPV	µm <sup>3</sup> ; fl	8.2	7.7			7.2					± 2.0
LYMPHO	#	3.38	3.44			3.33					± 1.00
	%	17.5	18.0			17.5					± 5.0
MONO	#	0.87	1.24			0.86					± 0.80
	%	4.5	6.5			4.5					± 4.0
GRANULO	#	15.05	14.41			14.82					± 1.40
	%	78.0	75.5			78.0					± 7.0