


LOT		NX 441		CONTROL		 (Exp.) 2023-07-05 (YYYY - MM - DD)							
Rev 1		ABX											
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA			PENTRA	PENTRA			PENTRA	PENTRA		
		DX120	DX NEXUS			DX120	DX NEXUS			DX120	DX NEXUS		
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	7.4	7.4		± 0.8	18.0	18.0		± 1.5	2.6	2.6		± 0.4
GBC CWBC	10 ³ /mm ³ ; 10 ⁹ /l	7.2	7.2		± 0.8	16.5	16.5		± 1.4	1.6	1.6		± 0.4
ERB	%	3.0	3.0		± 1.7	8.8	8.8		± 3.3	65.2	65.2		± 23.0
ERB	#	0.216	0.216		± 0.100	1.452	1.452		± 0.700	1.043	1.043		± 0.300

* les valeurs sont données à titre d'indication et doivent être utilisées comme RuO (Utilisation en Recherche Seulement) Indicative expected values, must be used as RuO (Research used only)