

 (Exp.) **2023-07-05**
(YYYY-MM-DD)

LOT		PX 441		Rev 2		CONTROL		ABX Lysebio											
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
			DX120 DF120	DX NEXUS DF NEXUS						DX120 DF120	DX NEXUS DF NEXUS						DX120 DF120		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.9	2.9			± 0.4	8.6	8.6			± 1.0	19.7	19.7			± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.35			± 0.16	4.61	4.61			± 0.20	5.14	5.14			± 0.25			
HB HGB	g/dl	6.1	6.1			± 0.4	13.7	13.7			± 0.5	16.2	16.2			± 0.6			
	g/l	61	61			± 4	137	137			± 5	162	162			± 6			
	mmol/l	3.79	3.79			± 0.25	8.51	8.51			± 0.31	10.06	10.06			± 0.37			
HT HCT	%	18.3	18.3			± 1.5	40.1	40.1			± 2.0	46.8	46.8			± 2.5			
	l/l	0.183	0.183			± 0.015	0.401	0.401			± 0.020	0.468	0.468			± 0.025			
VGM MCV	µm ³ ; fl	78	78			± 5	87	87			± 5	91	91			± 5			
TGMH MCH	pg	26.0	26.0			± 2.0	29.7	29.7			± 2.0	31.5	31.5			± 2.5			
	fmol	1.61	1.61			± 0.12	1.85	1.85			± 0.12	1.96	1.96			± 0.16			
CCMH MCHC	g/dl	33.3	33.3			± 3.0	34.2	34.2			± 3.0	34.6	34.6			± 3.0			
	g/l	333	333			± 30	342	342			± 30	346	346			± 30			
	mmol/l	20.67	20.67			± 1.86	21.21	21.21			± 1.86	21.51	21.51			± 1.86			
IDR RDW	%	17.0	17.0			± 4.0	15.0	15.0			± 4.0	13.5	13.5			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	76	76			± 20	249	249			± 30	492	492			± 50			
VPM MPV	µm ³ ; fl	8.9	8.9			± 2.0	8.6	8.6			± 2.0	8.7	8.7			± 2.0			
NEUT	#	1.56	1.56			± 0.35	4.88	4.88			± 0.90	14.60	14.60			± 1.90			
	%	53.7	53.7			± 10.0	56.7	56.7			± 10.0	74.2	74.2			± 10.0			
LYMPHO	#	0.95	0.95			± 0.33	2.89	2.89			± 0.70	2.92	2.92			± 1.50			
	%	32.6	32.6			± 12.0	33.6	33.6			± 8.0	14.8	14.8			± 8.0			
MONO	#	0.08	0.08			± 0.08	0.17	0.17			± 0.17	0.43	0.43			± 0.43			
	%	2.7	2.7			± 2.7	2.0	2.0			± 2.0	2.2	2.2			± 2.2			
EOS	#	0.21	0.21			± 0.21	0.28	0.28			± 0.28	0.91	0.91			± 0.91			
	%	7.2	7.2			± 7.2	3.3	3.3			± 3.3	4.6	4.6			± 4.6			
BASO	#	0.11	0.11			± 0.11	0.38	0.38			± 0.38	0.83	0.83			± 0.83			
	%	3.8	3.8			± 3.8	4.4	4.4			± 4.4	4.2	4.2			± 4.2			

1300090306
BACK / VERSO
Ref. TEMP-0821 Rev.47