

ABX Minotrol Retic



LOT RX 441
Rev 1

CONTROL

(Exp.) 2023-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.3	1.3	± 0.7	6.2	6.2	± 2.5	10.0	10.0	± 3.0			
RETIC	%	1.46	1.46	± 0.73	4.84	4.84	± 1.21	9.78	9.78	± 1.96			
RETIC	10 ⁶ /µl	0.051	0.051	± 0.026	0.168	0.168	± 0.042	0.333	0.333	± 0.067			
MFI MRI *	%	29.7	29.7	± 8.0	46.4	46.4	± 10.0	52.3	52.3	± 12.0			
GR RBC	10 ⁶ /µl	3.47	3.47	± 0.30	3.48	3.48	± 0.30	3.41	3.41	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.42 FRONT / RECTO 1300090305

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 441
Rev 1

CONTROL

(Exp.) 2023-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
XLR			XLR			XLR							
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	6.2			± 2.5	10.0			± 3.0
RETIC	%	1.70			± 0.85	5.38			± 1.35	11.19			± 2.24
RETIC	10 ⁶ /µl	0.059			± 0.030	0.188			± 0.047	0.382			± 0.076
MFI MRI *	%	30.0			± 8.0	40.1			± 10.0	43.3			± 12.0
GR RBC	10 ⁶ /µl	3.47			± 0.30	3.49			± 0.30	3.41			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.42 BACK / VERSO 1300090305

* For research use only in Canada and USA