


LOT PX 441
Rev 2


CONTROL

 (Exp.) 2023-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA					PENTRA					PENTRA					
			XL80	XLR				XL80	XLR				XL80	XLR				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.9	2.9			± 0.4	8.4	8.4			± 1.0	19.0	19.0			± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.27	2.27			± 0.16	4.52	4.52			± 0.20	5.00	5.00			± 0.25	
HB	HGB	g/dl	5.9	5.9			± 0.4	13.7	13.7			± 0.5	16.3	16.3			± 0.6	
		g/l	59	59			± 4	137	137			± 5	163	163			± 6	
		mmol/l	3.66	3.66			± 0.25	8.51	8.51			± 0.31	10.12	10.12			± 0.37	
HT	HCT	%	17.7	17.7			± 1.5	39.3	39.3			± 2.0	45.5	45.5			± 2.5	
		l/l	0.177	0.177			± 0.015	0.393	0.393			± 0.020	0.455	0.455			± 0.025	
VGM	MCV	µm ³ ; fl	78	78			± 5	87	87			± 5	91	91			± 5	
TGMH	MCH	pg	26.0	26.0			± 2.0	30.3	30.3			± 2.0	32.6	32.6			± 2.5	
		fmol	1.61	1.61			± 0.12	1.88	1.88			± 0.12	2.02	2.02			± 0.16	
CCMH	MCHC	g/dl	33.3	33.3			± 3.0	34.8	34.8			± 3.0	35.8	35.8			± 3.0	
		g/l	333	333			± 30	348	348			± 30	358	358			± 30	
		mmol/l	20.69	20.69			± 1.86	21.63	21.63			± 1.86	22.25	22.25			± 1.86	
IDR	RDW	%	14.5	14.5			± 4.0	13.0	13.0			± 4.0	12.5	12.5			± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72			± 20	246	246			± 30	487	487			± 50	
VPM	MPV	µm ³ ; fl	8.7	8.7			± 2.0	8.7	8.7			± 2.0	8.9	8.9			± 2.0	
NEUT		#	1.39	1.39			± 0.35	4.38	4.38			± 0.90	13.62	13.62			± 1.90	
		%	48.1	48.1			± 10.0	52.2	52.2			± 10.0	71.7	71.7			± 10.0	
LYMPHO		#	1.19	1.19			± 0.33	3.35	3.35			± 0.70	3.36	3.36			± 1.50	
		%	41.0	41.0			± 12.0	39.9	39.9			± 8.0	17.7	17.7			± 8.0	
MONO		#	0.06	0.06			± 0.06	0.15	0.15			± 0.15	0.42	0.42			± 0.42	
		%	2.0	2.0			± 2.0	1.8	1.8			± 1.8	2.2	2.2			± 2.2	
EOS		#	0.18	0.18			± 0.18	0.28	0.28			± 0.28	0.84	0.84			± 0.84	
		%	6.3	6.3			± 6.3	3.3	3.3			± 3.3	4.4	4.4			± 4.4	
BASO		#	0.08	0.08			± 0.08	0.24	0.24			± 0.24	0.76	0.76			± 0.76	
		%	2.6	2.6			± 2.6	2.8	2.8			± 2.8	4.0	4.0			± 4.0	

LOT PX 441
Rev 2

CONTROL

 (Exp.) 2023-07-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.9	2.9	2.9		± 0.4	8.4	8.4	8.3		± 1.0	19.4	19.2	19.1		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.37	2.33	2.32		± 0.16	4.56	4.54	4.52		± 0.20	5.02	4.98	5.00		± 0.25			
HB HGB	g/dl	6.0	6.0	6.0		± 0.4	13.7	13.6	13.7		± 0.5	16.2	16.2	16.2		± 0.6			
	g/l	60	60	60		± 4	137	136	137		± 5	162	162	162		± 6			
	mmol/l	3.73	3.73	3.73		± 0.25	8.51	8.45	8.51		± 0.31	10.06	10.06	10.06		± 0.37			
HT HCT	%	17.8	17.5	17.4		± 1.5	39.2	38.6	38.9		± 2.0	46.2	45.3	45.8		± 2.5			
	l/l	0.178	0.175	0.174		± 0.015	0.392	0.386	0.389		± 0.020	0.462	0.453	0.458		± 0.025			
VGM MCV	µm ³ ; fl	75	75	75.0		± 5	86	85	86.0		± 5	92	91	91.5		± 5			
TGMH MCH	pg	25.3	25.8	25.9		± 2.0	30.0	30.0	30.3		± 2.0	32.3	32.5	32.4		± 2.5			
	fmol	1.57	1.60	1.61		± 0.12	1.87	1.86	1.88		± 0.12	2.00	2.02	2.01		± 0.16			
CCMH MCHC	g/dl	33.8	34.3	34.5		± 3.0	34.9	35.2	35.2		± 3.0	35.1	35.7	35.4		± 3.0			
	g/l	338	343	345		± 30	349	352	352		± 30	351	357	354		± 30			
	mmol/l	20.96	21.32	21.41		± 1.86	21.69	21.89	21.89		± 1.86	21.78	22.20	21.99		± 1.86			
IDR RDW	%	15.0	15.0	13.5		± 4.0	13.5	13.0	12.0		± 4.0	12.0	12.0	11.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	71	73		± 20	244	242	237		± 30	485	485	466		± 50			
VPM MPV	µm ³ ; fl	8.6	8.6	8.0		± 2.0	8.1	8.2	8.0		± 2.0	8.1	8.3	7.9		± 2.0			
NEUT	#	1.39	1.39	1.42		± 0.35	4.35	4.37	4.36		± 0.90	13.81	13.77	13.75		± 1.90			
	%	47.9	48.1	49.0		± 10.0	51.8	52.0	52.5		± 10.0	71.2	71.7	72.0		± 10.0			
LYMPHO	#	1.19	1.19	1.16		± 0.33	3.38	3.37	3.32		± 0.70	3.51	3.49	3.44		± 1.50			
	%	41.1	41.2	40.0		± 12.0	40.2	40.1	40.0		± 8.0	18.1	18.2	18.0		± 8.0			
MONO	#	0.07	0.05	0.06		± 0.05	0.17	0.15	0.12		± 0.12	0.45	0.33	0.29		± 0.29			
	%	2.4	1.7	2.0		± 1.7	2.0	1.8	1.5		± 1.5	2.3	1.7	1.5		± 1.5			
EOS	#	0.17	0.18	0.17		± 0.17	0.26	0.27	0.25		± 0.25	0.85	0.84	0.86		± 0.84			
	%	6.0	6.3	6.0		± 6.0	3.1	3.2	3.0		± 3.0	4.4	4.4	4.5		± 4.4			
BASO	#	0.08	0.08	0.09		± 0.08	0.24	0.24	0.25		± 0.24	0.78	0.77	0.76		± 0.76			
	%	2.6	2.7	3.0		± 2.6	2.9	2.9	3.0		± 2.9	4.0	4.0	4.0		± 4.0			

1300090306 FRONT / RECTO Ref: TEMP-0821 Rev.47