


ABX Minotrol Retic

LOT		RX 441				CONTROL				<div><div></div><div>(Exp.) 2023-07-05</div><div>(YYYY - MM - DD)</div></div>						
		Rev 1		ABX												
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			1	TOLERANCES TOLERANCE	CONTROL			2	TOLERANCES TOLERANCE	CONTROL			3	TOLERANCES TOLERANCE
			PENTRA	PENTRA			PENTRA	PENTRA		PENTRA		PENTRA				
			DX120	DX Nexus			DX120	DX Nexus		DX120		DX Nexus				
Méthode Manuelle / Manual RETIC	%		1.3	1.3	± 0.7		6.2	6.2	± 2.5		10.0	10.0	± 3.0			
RETIC	%		1.46	1.46	± 0.73		4.84	4.84	± 1.21		9.78	9.78	± 1.96			
RETIC	10 ⁶ /µl		0.051	0.051	± 0.026		0.168	0.168	± 0.042		0.333	0.333	± 0.067			
MFI MRI *	%		29.7	29.7	± 8.0		46.4	46.4	± 10.0		52.3	52.3	± 12.0			
GR RBC	10 ⁶ /µl		3.47	3.47	± 0.30		3.48	3.48	± 0.30		3.41	3.41	± 0.30			
PIC RANGE	-		27.0	27.0	± 5.0		27.0	27.0	± 5.0		27.0	27.0	± 5.0			

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.42 FRONT / RECTO 1300090305

ABX Minotrol Retic

LOT		RX 441		CONTROL		 (Exp.) 2023-07-05 (YYYY - MM - DD)										
Rev 1		ABX														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			1	TOLERANCES TOLERANCE	CONTROL			2	TOLERANCES TOLERANCE	CONTROL			3	TOLERANCES TOLERANCE
		PENTRA			PENTRA				PENTRA							
		XLR			XLR				XLR							
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	6.2			± 2.5	10.0			± 3.0			
RETIC	%	1.70			± 0.85	5.38			± 1.35	11.19			± 2.24			
RETIC	10 ⁶ /µl	0.059			± 0.030	0.188			± 0.047	0.382			± 0.076			
MFI MRI *	%	30.0			± 8.0	40.1			± 10.0	43.3			± 12.0			
GR RBC	10 ⁶ /µl	3.47			± 0.30	3.49			± 0.30	3.41			± 0.30			
PIC RANGE	-	8			± 4	8			± 4	8			± 4			

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.42 BACK / VERSO 1300090305