


LOT PX 441
Rev 2

CONTROL

 (Exp.) 2023-07-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Whitediff | | | | | | | | | | | | TOLERANCES TOLERANCE | | | |
|--------------------------|--|----------------------|--------------------|--------------------|--------------------|-------------------------|----------------------|--------------------|--------------------|--------------------|-------------------------|----------------------|--------------------|-------------------------|--------------------|-------------------------|--------------------|
| | | CONTROL | | | | TOLERANCES TOLERANCE | CONTROL | | | | TOLERANCES TOLERANCE | CONTROL | | | | TOLERANCES TOLERANCE | |
| | | YUMIZEN | | | | | YUMIZEN | | | | | YUMIZEN | | | | | |
| | | H550 V1.0 to V2.x | H500 CT H500 CT | H500 OT H500 OT | H500 CT H500 CT | | H550 V1.0 to V2.x | H500 CT H500 CT | H500 OT H500 OT | H500 CT H500 CT | | H550 V1.0 to V2.x | H500 CT H500 CT | | H500 OT H500 OT | | H500 CT H500 CT |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.90 | 2.90 | 2.90 | 2.90 | ± 0.40 | 8.20 | 8.20 | 8.20 | 8.20 | ± 1.00 | 18.40 | 18.40 | 18.40 | 18.40 | ± 2.20 | |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.32 | 2.32 | 2.32 | 2.32 | ± 0.16 | 4.59 | 4.59 | 4.64 | 4.64 | ± 0.20 | 5.11 | 5.11 | 5.11 | 5.11 | ± 0.25 | |
| HB HGB | g/dl | 5.8 | 5.8 | 5.9 | 5.9 | ± 0.4 | 13.2 | 13.2 | 13.3 | 13.3 | ± 0.5 | 15.7 | 15.7 | 15.7 | 15.7 | ± 0.6 | |
| | g/l | 58 | 58 | 59 | 59 | ± 4 | 132 | 132 | 133 | 133 | ± 5 | 157 | 157 | 157 | 157 | ± 6 | |
| | mmol/l | 3.60 | 3.60 | 3.66 | 3.66 | ± 0.25 | 8.20 | 8.20 | 8.26 | 8.26 | ± 0.31 | 9.75 | 9.75 | 9.75 | 9.75 | ± 0.37 | |
| HT HCT | % | 17.7 | 17.7 | 17.6 | 17.1 | ± 1.5 | 39.9 | 39.9 | 38.8 | 37.8 | ± 2.0 | 47.0 | 47.0 | 46.3 | 44.8 | ± 2.5 | |
| | l/l | 0.177 | 0.177 | 0.176 | 0.171 | ± 0.015 | 0.399 | 0.399 | 0.388 | 0.378 | ± 0.020 | 0.470 | 0.470 | 0.463 | 0.448 | ± 0.025 | |
| VGM MCV | µm ³ ; fl | 76.5 | 76.5 | 75.9 | 73.7 | ± 5.0 | 87.0 | 87.0 | 83.6 | 81.5 | ± 5.0 | 92.0 | 92.0 | 90.6 | 87.7 | ± 5.0 | |
| TGMH MCH | pg | 25.0 | 25.0 | 25.4 | 25.4 | ± 2.0 | 28.8 | 28.8 | 28.7 | 28.7 | ± 2.0 | 30.7 | 30.7 | 30.7 | 30.7 | ± 2.5 | |
| | fmol | 1.55 | 1.55 | 1.58 | 1.58 | ± 0.12 | 1.79 | 1.79 | 1.78 | 1.78 | ± 0.12 | 1.91 | 1.91 | 1.91 | 1.91 | ± 0.16 | |
| CCMH MCHC | g/dl | 32.7 | 32.7 | 33.5 | 34.5 | ± 3.0 | 33.1 | 33.1 | 34.3 | 35.2 | ± 3.0 | 33.4 | 33.4 | 33.9 | 35.0 | ± 3.0 | |
| | g/l | 327 | 327 | 335 | 345 | ± 30 | 331 | 331 | 343 | 352 | ± 30 | 334 | 334 | 339 | 350 | ± 30 | |
| | mmol/l | 20.31 | 20.31 | 20.80 | 21.42 | ± 1.86 | 20.56 | 20.56 | 21.30 | 21.86 | ± 1.86 | 20.74 | 20.74 | 21.05 | 21.74 | ± 1.86 | |
| IDR-SD RDW-SD | fl | 45.0 | 45.0 | 35.4 | 35.4 | ± 8.0 | 46.0 | 46.0 | 37.8 | 37.8 | ± 8.0 | 45.0 | 45.0 | 37.5 | 37.5 | ± 8.0 | |
| IDR-CV RDW-CV | % | 16.0 | 16.0 | 15.0 | 15.0 | ± 4.0 | 15.0 | 15.0 | 13.3 | 13.3 | ± 4.0 | 14.0 | 14.0 | 12.5 | 12.5 | ± 4.0 | |
| PLA. PLT | 10 ⁹ /mm ³ ; 10 ⁹ /l | 68 | 68 | 82 | 85 | ± 20 | 240 | 240 | 245 | 257 | ± 30 | 489 | 489 | 486 | 510 | ± 50 | |
| VMP MPV | µm ³ ; fl | 9.1 | 9.1 | 10.2 | 10.2 | ± 2.0 | 9.3 | 9.3 | 9.0 | 9.0 | ± 2.0 | 9.3 | 9.3 | 8.9 | 8.9 | ± 2.0 | |
| NEU | 10 ⁹ /mm ³ ; 10 ⁹ /l | 1.19 | 1.19 | 1.14 | 1.14 | ± 0.35 | 3.87 | 3.87 | 3.72 | 3.72 | ± 0.90 | 12.57 | 12.57 | 12.27 | 12.27 | ± 1.90 | |
| | % | 41.2 | 41.2 | 39.4 | 39.4 | ± 10.0 | 47.2 | 47.2 | 45.4 | 45.4 | ± 10.0 | 68.3 | 68.3 | 66.7 | 66.7 | ± 10.0 | |
| LYM | 10 ⁹ /mm ³ ; 10 ⁹ /l | 1.21 | 1.21 | 1.25 | 1.25 | ± 0.33 | 3.44 | 3.44 | 3.54 | 3.54 | ± 0.70 | 3.62 | 3.62 | 3.75 | 3.75 | ± 1.50 | |
| | % | 41.7 | 41.7 | 43.0 | 43.0 | ± 12.0 | 42.0 | 42.0 | 43.2 | 43.2 | ± 8.0 | 19.7 | 19.7 | 20.4 | 20.4 | ± 8.0 | |
| MON | 10 ⁹ /mm ³ ; 10 ⁹ /l | 0.21 | 0.21 | 0.22 | 0.22 | ± 0.21 | 0.40 | 0.40 | 0.44 | 0.44 | ± 0.40 | 0.74 | 0.74 | 0.83 | 0.83 | ± 0.74 | |
| | % | 7.2 | 7.2 | 7.6 | 7.6 | ± 7.2 | 4.9 | 4.9 | 5.4 | 5.4 | ± 4.9 | 4.0 | 4.0 | 4.5 | 4.5 | ± 4.0 | |
| EOS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 0.18 | 0.18 | 0.18 | 0.18 | ± 0.18 | 0.24 | 0.24 | 0.25 | 0.25 | ± 0.24 | 0.72 | 0.72 | 0.79 | 0.79 | ± 0.72 | |
| | % | 6.1 | 6.1 | 6.2 | 6.2 | ± 6.1 | 2.9 | 2.9 | 3.0 | 3.0 | ± 2.9 | 3.9 | 3.9 | 4.3 | 4.3 | ± 3.9 | |
| BAS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 0.11 | 0.11 | 0.11 | 0.11 | ± 0.11 | 0.25 | 0.25 | 0.25 | 0.25 | ± 0.25 | 0.75 | 0.75 | 0.75 | 0.75 | ± 0.75 | |
| | % | 3.8 | 3.8 | 3.8 | 3.8 | ± 3.8 | 3.0 | 3.0 | 3.0 | 3.0 | ± 3.0 | 4.1 | 4.1 | 4.1 | 4.1 | ± 4.1 | |
| IMG | 10 ⁹ /mm ³ ; 10 ⁹ /l | N/A | N/A | 0.09 | 0.09 | ± 0.09 | N/A | N/A | 0.32 | 0.32 | ± 0.32 | N/A | N/A | 0.96 | 0.96 | ± 0.96 | |
| | % | N/A | N/A | 3.2 | 3.2 | ± 3.2 | N/A | N/A | 3.9 | 3.9 | ± 3.9 | N/A | N/A | 5.2 | 5.2 | ± 5.2 | |

Ref: TEMP-0821 Rev.47 FRONT / RECTO 1300105209