

ABX Minocal


 (Exp.) 2023-06-05
 (YYYY - MM - DD)

LOT		CX 482		CAL		(Exp.) 2023-06-05 (YYYY - MM - DD)					
		Rev 1									
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60	DX120 - DF120	60 60 C+ ES60	XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.2	9.4	9.3	9.1	9.4	9.2	9.4	9.2	9.1	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.39	4.38	4.40	4.35	4.41	4.34	4.41	4.38	4.35	± 0.06
HB HGB	g/dl	13.1	13.0	13.1	13.2	13.1	13.1	13.1	13.2	13.2	± 0.2
	g/l	131	130	131	132	131	131	131	132	132	± 2
	mmol/l	8.14	8.07	8.14	8.20	8.14	8.14	8.14	8.20	8.20	± 0.12
HT HCT	%	36.9	36.8	35.9	36.1	35.9	35.6	35.9	35.9	36.1	± 1.0
	l/l	0.369	0.368	0.359	0.361	0.359	0.356	0.359	0.359	0.361	± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	253	247	238	243	253	249	253	249	243	± 10
VMP MPV	µm ³ ; fl	8.5	8.5	N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.48 BACK / VERSO 1300090308