



Yumizen G DDi 2 for YUMIZEN G1500/G1550



Reference No.:		1300036391	
Lot No.:		930115	
Expiry date:		2024-01-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G1500/G1550	µg(FEU)/mL	ΔOD
		4,00	0,2740
		3,00	0,2030
		2,00	0,1240
		1,00	0,0550
		0,50	0,0260
		0,25	0,0130
Slope value		0,06	

Yumizen G DDi 2 for YUMIZEN G800



Reference No.:		1300036391	
Lot No.:		930115	
Expiry date:		2024-01-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G800	µg(FEU)/mL	ΔOD
		4,00	0,2620
		3,00	0,2015
		2,00	0,1230
		1,00	0,0515
		0,50	0,0235
		0,25	0,0110
Slope value		0,06	



Yumizen G DDi 2 ***for YUMIZEN*** ***G400 DDi/G200***

Reference No.:		1300036391	
Lot No.:		930115	
Expiry date:		2024-01-31	
Coagulometers		MASTER CURVE	
		µg(FEU)/mL	ΔOD
Optical	Yumizen G400 DDi/G200	4,00	0,1135
		3,00	0,0830
		2,00	0,0505
		1,00	0,0223
		0,50	0,0107
		0,25	0,0052
		Slope value	0,02
Important notice			
<p>For Yumizen G400 DDi and Yumizen G200, make sure that the end time is set to 180 seconds in the D-Dimer setup and the curve type Lin-Lin p-p is selected in Calibration > D-Dimer > ug FEU/ml. For more information, please contact your local HORIBA Medical representative.</p>			