



Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 2 0 6 2 8 4 6 4 1 2 0 6 2 4 D



D 5 1 0 9 1 4 8 0 7 8 1 3 0 0 7 9 D



D 6 1 1 8 1 0 5 1 4 2 0 6 9 1 1 6 D



D 7 0 8 3 1 2 5 0 8 5 1 1 5 0 7 9 D



D 8 1 3 1 0 7 9 1 1 8 3 3 0 4 4 6 D



D 9 2 8 9 3 9 1 2 0 0 3 0 0 1 9 6 D



D 0 2 9 4 3 7 3 D

Reference No.:		1300036412	
Lot No:		920628 (rev. 1)	
Expiry date:		2024-06-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,9-14,8
	Prothrombin %	%	78-130
	Prothrombin INR	INR	0,79-1,18
PT Liq	Prothrombin Time	sec	10,5-14,2
	Prothrombin %	%	69-116
	Prothrombin INR	INR	0,83-1,25
PT R	Prothrombin Time	sec	8,5-11,5
	Prothrombin %	%	79-131
	Prothrombin INR	INR	0,79-1,18
APTT	Partial Thromboplastin Time	sec	33,0-44,6
APTT Liq	Partial Thromboplastin Time	sec	28,9-39,1
FIB Cl	Fibrinogen	g/L	2,0-3,0
TT	Thrombin Time	sec	19,6-29,4
AT	Antithrombin	%	80-108

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 2 0 6 2 8 5 6 4 1 2 0 6 2 4 D



D 5 1 6 2 2 4 3 0 3 0 0 5 0 1 3 7 D



D 6 2 0 5 1 6 3 2 4 4 0 2 5 0 4 2 D



D 7 1 5 2 2 2 7 1 2 0 1 8 0 0 3 4 D



D 8 0 5 6 1 1 8 1 7 6 6 1 8 8 3 6 D



D 9 6 0 0 8 1 2 0 6 0 1 5 0 0 0 0 D



D 0 9 9 9 3 3 4 D

Reference No.:		1300036412	
Lot No:		920628 (rev. 1)	
Expiry date:		2024-06-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	16,2-24,3
	Prothrombin %	%	30-50
	Prothrombin INR	INR	1,37-2,05
PT Liq	Prothrombin Time	sec	16,3-24,4
	Prothrombin %	%	25-42
	Prothrombin INR	INR	1,52-2,27
PT R	Prothrombin Time	sec	12,0-18,0
	Prothrombin %	%	34-56
	Prothrombin INR	INR	1,18-1,76
APTT	Partial Thromboplastin Time	sec	61,8-83,6
APTT Liq	Partial Thromboplastin Time	sec	60-81,2
FIB Cl	Fibrinogen	g/L	0,6-1,5
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	32-54
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			