

Yumizen G PT Reco 5 for Yumizen G Line



Reference No.:		1300036375	
Lot No.:		920827(rev. 1)	
Expiry date:		2024-08-31	
Coagulometers		ISI	MASTER CURVE
Optical	Yumizen G Line		%
		100	10,7
		50	16,1
		33	21,6
		25	26,8
	0,94	MNPT	10,7
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>For Yumizen G200, Yumizen G400 and Yumizen G400 DDi, the Min. Time must be set to 7 seconds and the Lag time to 5 seconds in the PT setup. Please refer to the Yumizen G PT Reco 5 reagent notice.</p>			