



# Yumizen G CTRL I (level 1) for Yumizen G Line



Reference No.:		1300036412	
Lot No:		920933 (rev. 1)	
Expiry date:		2024-09-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,9-14,8
	Prothrombin %	%	71-118
	Prothrombin INR	INR	0,82-1,23
PT Liq	Prothrombin Time	sec	12,3-16,6
	Prothrombin %	%	64-106
	Prothrombin INR	INR	0,88-1,31
PT R	Prothrombin Time	sec	8,8-11,9
	Prothrombin %	%	73-122
	Prothrombin INR	INR	0,81-1,21
APTT	Partial Thromboplastin Time	sec	34,7-47,0
APTT Liq	Partial Thromboplastin Time	sec	27,6-37,3
FIB Cl	Fibrinogen	g/L	2,0-3,0
TT	Thrombin Time	sec	22,6-33,9
AT	Antithrombin	%	92-125

### Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.  
 The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.



# Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 2 0 9 3 3 5 6 4 1 2 0 9 2 4 D



D 5 1 3 9 2 0 8 0 3 4 0 5 7 1 1 6 D



D 6 1 7 4 1 9 2 2 8 8 0 2 7 0 4 5 D



D 7 1 6 1 2 4 2 1 1 6 1 7 5 0 4 0 D



D 8 0 6 7 1 1 3 1 6 9 5 6 7 7 6 7 D



D 9 5 7 8 7 8 2 0 7 0 1 5 0 0 0 0 D



D 0 9 9 9 3 9 8 D

Reference No.:		1300036412	
Lot No:		920933 (rev. 1)	
Expiry date:		2024-09-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	13,9-20,8
	Prothrombin %	%	34-57
	Prothrombin INR	INR	1,16-1,74
PT Liq	Prothrombin Time	sec	19,2-28,8
	Prothrombin %	%	27-45
	Prothrombin INR	INR	1,61-2,42
PT R	Prothrombin Time	sec	11,6-17,5
	Prothrombin %	%	40-67
	Prothrombin INR	INR	1,13-1,69
APTT	Partial Thromboplastin Time	sec	56,7-76,7
APTT Liq	Partial Thromboplastin Time	sec	57,8-78,2
FIB Cl	Fibrinogen	g/L	0,7-1,5
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	29-48
<b>Important notice</b>			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			