

# ABX Minocal



**LOT**

**CX 483**

**CAL**



**(Exp.) 2023-07-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60	DX120 - DF120	60 60 C+ ES60	XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	9.1	9.1	9.3	9.0	9.3	9.0	9.3	9.1	9.0	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	4.46	4.44	4.45	4.47	4.53	4.48	4.53	4.50	4.47	± 0.06
HB HGB	g/dl	13.1	13.0	13.1	13.2	13.2	13.2	13.2	13.2	13.2	± 0.2
	g/l	131	130	131	132	132	132	132	132	132	± 2
	mmol/l	8.14	8.07	8.14	8.20	8.20	8.20	8.20	8.20	8.20	± 0.12
HT HCT	%	36.8	36.6	35.6	35.8	36.0	35.8	36.0	36.0	35.8	± 1.0
	l/l	0.368	0.366	0.356	0.358	0.360	0.358	0.360	0.360	0.358	± 0.010
PLA PLT	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	244	246	240	236	250	246	250	244	236	± 10
VMP MPV	µm <sup>3</sup> ; fl	8.4	8.3	N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.49 BACK / VERSO 1300090308