

LOT MX 442 L
Rev 1

CONTROL L

 (Exp.)

2023-10-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	1.9			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.23	2.25			2.19					± 0.17
HB HGB	g/dl	5.7	5.8			5.6					± 0.5
	g/l	57	58			56					± 5
	mmol/l	3.54	3.59			3.48					± 0.31
HT HCT	%	15.6	15.8			15.5					± 2.0
	l/l	0.156	0.158			0.155					± 0.020
VGM MCV	µm ³ ; fl	70	70			71					± 4
TGMH MCH	pg	25.6	25.7			25.6					± 2.0
	fmol	1.59	1.60			1.59					± 0.12
CCMH MCHC	g/dl	36.5	36.7			36.0					± 3.0
	g/l	365	367			360					± 30
	mmol/l	22.68	22.80			22.37					± 1.86
IDR RDW	%	15.0	14.0			15.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	79			80					± 22
VPM MPV	µm ³ ; fl	8.4	8.2			7.5					± 2.0
LYMPHO	#	1.26	1.22			1.26					± 0.40
	%	64.5	65.1			66.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.5	9.5			7.0					± 6.0
GRANULO	#	0.54	0.48			0.54					± 0.40
	%	27.0	25.5			27.0					± 7.0

LOT **MX 442** **N**
 Rev 1

CONTROL **N**

(Exp.)

2023-10-05
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.0	7.2			7.1					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.65			4.54					± 0.26
HB HGB	g/dl	13.4	13.5			13.3					± 0.7
	g/l	134	135			133					± 7
	mmol/l	8.32	8.38			8.28					± 0.43
HT HCT	%	37.3	37.2			37.7					± 2.7
	l/l	0.373	0.372			0.377					± 0.027
VGM MCV	µm ³ ; fl	81	80			83					± 4
TGMH MCH	pg	29.1	29.0			29.4					± 2.0
	fmol	1.81	1.80			1.82					± 0.12
CCMH MCHC	g/dl	36.0	36.3			35.4					± 3.0
	g/l	360	363			354					± 30
	mmol/l	22.33	22.54			21.96					± 1.86
IDR RDW	%	14.5	14.5			15.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	244	260			260					± 40
VPM MPV	µm ³ ; fl	8.2	7.8			7.1					± 2.0
LYMPHO	#	2.55	2.75			2.56					± 0.40
	%	36.5	38.2			36.0					± 5.0
MONO	#	0.49	0.54			0.46					± 0.30
	%	7.0	7.6			6.5					± 4.0
GRANULO	#	3.96	3.92			4.08					± 0.60
	%	56.5	54.4			57.5					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143



LOT

MX 442 H
 Rev 1

CONTROL H

(Exp.)

2023-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	19.6	19.7			19.4					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.65	5.68			5.61					± 0.32
HB HGB	g/dl	18.2	18.0			18.0					± 0.9
	g/l	182	180			180					± 9
	mmol/l	11.30	11.18			11.19					± 0.56
HT HCT	%	49.7	50.0			50.5					± 3.4
	l/l	0.497	0.500			0.505					± 0.034
VGM MCV	µm ³ ; fl	88	88			90					± 4
TGMH MCH	pg	32.2	31.7			32.1					± 2.0
	fmol	2.00	1.97			1.99					± 0.12
CCMH MCHC	g/dl	36.6	36.0			35.7					± 3.0
	g/l	366	360			357					± 30
	mmol/l	22.73	22.36			22.16					± 1.86
IDR RDW	%	14.0	14.0			15.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	526	560			559					± 65
VPM MPV	µm ³ ; fl	8.0	7.5			7.2					± 2.0
LYMPHO	#	3.33	3.55			3.30					± 1.00
	%	17.0	18.0			17.0					± 5.0
MONO	#	0.98	1.49			0.87					± 0.80
	%	5.0	7.6			4.5					± 4.0
GRANULO	#	15.29	14.67			15.23					± 1.40
	%	78.0	74.5			78.5					± 7.0