

# ABX Minotrol CRP



**LOT**

**MC 428**

**CONTROL**



(Exp.) **2023-10-05**  
(YYYY-MM-DD)

Rev 1

		ABX Lyse														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	2.0	2.0			± 0.4	7.2	7.3			± 1.0	19.3	19.8			± 2.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	2.31	2.30			± 0.15	4.65	4.69			± 0.18	5.58	5.64			± 0.25
HB HGB	g/dl	5.9	6.0			± 0.4	13.5	13.7			± 0.5	17.8	17.9			± 0.7
	g/l	59	60			± 4	135	137			± 5	178	179			± 7
	mmol/l	3.66	3.73			± 0.25	8.38	8.51			± 0.31	11.05	11.12			± 0.44
HT HCT	%	15.9	16.1			± 2.0	37.2	37.5			± 2.5	49.7	50.2			± 3.0
	l/l	0.159	0.161			± 0.020	0.372	0.375			± 0.025	0.497	0.502			± 0.030
VGM MCV	µm <sup>3</sup> ; fl	69	70			± 4	80	80			± 4	89	89			± 4
TGMH MCH	pg	25.5	26.1			± 2.4	29.0	29.2			± 2.4	31.9	31.7			± 2.8
	fmol	1.59	1.62			± 0.15	1.80	1.81			± 0.15	1.98	1.97			± 0.18
CCMH MCHC	g/dl	37.0	37.3			± 3.0	36.3	36.5			± 3.0	35.8	35.7			± 3.0
	g/l	370	373			± 30	363	365			± 30	358	357			± 30
	mmol/l	22.99	23.14			± 1.86	22.54	22.68			± 1.86	22.26	22.14			± 1.86
IDR RDW	%	14.0	14.0			± 3.0	13.5	13.0			± 3.0	13.0	12.5			± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	84	80			± 20	268	265			± 40	559	552			± 60
VPM MPV	µm <sup>3</sup> ; fl	8.5	7.6			± 2.0	8.0	7.3			± 2.0	7.7	7.2			± 2.0
LYMPHO	#	1.3	1.3			± 0.2	2.6	2.6			± 0.5	3.2	3.3			± 1.1
	%	64.5	63.5			± 8.0	36.0	35.5			± 6.0	16.5	16.5			± 5.0
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.7	1.0			± 1.0
	%	10.5	9.5			± 7.0	8.5	7.5			± 5.0	9.0	5.0			± 5.0
GRANULO	#	0.5	0.5			± 0.2	4.0	4.2			± 0.5	14.4	15.5			± 1.5
	%	25.0	27.0			± 7.0	55.5	57.0			± 6.0	74.5	78.5			± 7.0
CRP	mg/dl	0.37	N/A			± 0.30	2.19	N/A			± 0.70	4.32	N/A			± 1.00
	mg/l	3.7	N/A			± 3.0	21.9	N/A			± 7.0	43.2	N/A			± 10.0

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170