

ABX Minotrol CRP



LOT

MC 428

CONTROL



(Exp.) **2023-10-05**
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0		± 0.4	7.2	7.3		± 1.0	18.9	19.7		± 2.2			
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.31	2.24		± 0.15	4.65	4.64		± 0.18	5.58	5.63		± 0.25			
HB	HGB	g/dl	5.6	5.7		± 0.4	13.2	13.6		± 0.5	17.8	18.3		± 0.7			
		g/l	56	57		± 4	132	136		± 5	178	183		± 7			
		mmol/l	3.48	3.54		± 0.25	8.22	8.45		± 0.31	11.05	11.36		± 0.44			
HT	HCT	%	15.9	16.1		± 2.0	37.2	38.0		± 2.5	49.7	50.7		± 3.0			
		l/l	0.159	0.161		± 0.020	0.372	0.380		± 0.025	0.497	0.507		± 0.030			
VGM	MCV	µm ³ ; fl	69	72		± 4	80	82		± 4	89	90		± 4			
TGMH	MCH	pg	24.3	25.4		± 2.4	28.5	29.3		± 2.4	31.9	32.5		± 2.8			
		fmol	1.51	1.58		± 0.15	1.77	1.82		± 0.15	1.98	2.02		± 0.18			
CCMH	MCHC	g/dl	35.2	35.3		± 3.0	35.6	35.7		± 3.0	35.8	36.1		± 3.0			
		g/l	352	353		± 30	356	357		± 30	358	361		± 30			
		mmol/l	21.84	21.95		± 1.86	22.09	22.20		± 1.86	22.26	22.43		± 1.86			
IDR	RDW	%	14.0	14.5		± 3.0	13.5	14.5		± 3.0	13.0	13.5		± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	84	80		± 20	268	262		± 40	559	552		± 60			
VPM	MPV	µm ³ ; fl	8.5	8.5		± 2.0	8.0	8.3		± 2.0	7.7	8.1		± 2.0			
LYMPHO		#	1.3	1.3		± 0.2	2.7	2.7		± 0.5	3.3	3.4		± 1.1			
		%	67.1	65.5		± 8.0	37.8	36.5		± 6.0	17.7	17.5		± 5.0			
MONO		#	0.2	0.2		± 0.2	0.6	0.5		± 0.4	1.7	0.9		± 0.9			
		%	8.4	7.5		± 7.0	8.1	6.5		± 5.0	9.0	4.5		± 4.5			
GRANULO		#	0.5	0.5		± 0.2	3.9	4.2		± 0.5	14.0	15.4		± 1.5			
		%	24.5	27.0		± 7.0	54.4	57.0		± 6.0	73.8	78.0		± 7.0			
CRP		mg/dl	0.37	N/A		± 0.30	2.19	N/A		± 0.70	4.32	N/A		± 1.00			
		mg/l	3.7	N/A		± 3.0	21.9	N/A		± 7.0	43.2	N/A		± 10.0			

Ref: TEMP-1070 Rev.42 1300027267-A