

ABX Minotrol CRP



LOT

MC 428

CONTROL



(Exp.) **2023-10-05**
(YYYY - MM - DD)

Rev 1

		ABX Lyse															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.2	7.3			± 1.0	19.3	19.8			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.31	2.30			± 0.15	4.65	4.69			± 0.18	5.58	5.64			± 0.25	
HB HGB	g/dl	5.9	6.0			± 0.4	13.5	13.7			± 0.5	17.8	17.9			± 0.7	
	g/l	59	60			± 4	135	137			± 5	178	179			± 7	
	mmol/l	3.66	3.73			± 0.25	8.38	8.51			± 0.31	11.05	11.12			± 0.44	
HT HCT	%	15.9	16.1			± 2.0	37.2	37.5			± 2.5	49.7	50.2			± 3.0	
	l/l	0.159	0.161			± 0.020	0.372	0.375			± 0.025	0.497	0.502			± 0.030	
VGM MCV	µm ³ ; fl	69	70			± 4	80	80			± 4	89	89			± 4	
TGMH MCH	pg	25.5	26.1			± 2.4	29.0	29.2			± 2.4	31.9	31.7			± 2.8	
	fmol	1.59	1.62			± 0.15	1.80	1.81			± 0.15	1.98	1.97			± 0.18	
CCMH MCHC	g/dl	37.0	37.3			± 3.0	36.3	36.5			± 3.0	35.8	35.7			± 3.0	
	g/l	370	373			± 30	363	365			± 30	358	357			± 30	
	mmol/l	22.99	23.14			± 1.86	22.54	22.68			± 1.86	22.26	22.14			± 1.86	
IDR RDW	%	14.0	14.0			± 3.0	13.5	13.0			± 3.0	13.0	12.5			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	84	80			± 20	268	265			± 40	559	552			± 60	
VPM MPV	µm ³ ; fl	8.5	7.6			± 2.0	8.0	7.3			± 2.0	7.7	7.2			± 2.0	
LYMPHO	#	1.3	1.3			± 0.2	2.6	2.6			± 0.5	3.2	3.3			± 1.1	
	%	64.5	63.5			± 8.0	36.0	35.5			± 6.0	16.5	16.5			± 5.0	
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.7	1.0			± 1.0	
	%	10.5	9.5			± 7.0	8.5	7.5			± 5.0	9.0	5.0			± 5.0	
GRANULO	#	0.5	0.5			± 0.2	4.0	4.2			± 0.5	14.4	15.5			± 1.5	
	%	25.0	27.0			± 7.0	55.5	57.0			± 6.0	74.5	78.5			± 7.0	
CRP	mg/dl	0.37	N/A			± 0.30	2.19	N/A			± 0.70	4.32	N/A			± 1.00	
	mg/l	3.7	N/A			± 3.0	21.9	N/A			± 7.0	43.2	N/A			± 10.0	

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170