

# ABX Minotrol Retic



**LOT** RX 442  
Rev 1

**CONTROL**

(Exp.) 2023-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA		PENTRA		PENTRA		PENTRA		PENTRA		PENTRA	
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%		<b>1.9</b>	<b>1.9</b>	± 1.0		<b>5.7</b>	<b>5.7</b>	± 2.3		<b>11.9</b>	<b>11.9</b>	± 3.6
RETIC	%		<b>1.29</b>	<b>1.29</b>	± 0.65		<b>4.98</b>	<b>4.98</b>	± 1.25		<b>10.16</b>	<b>10.16</b>	± 2.03
RETIC	10 <sup>6</sup> /µl		<b>0.044</b>	<b>0.044</b>	± 0.022		<b>0.168</b>	<b>0.168</b>	± 0.042		<b>0.345</b>	<b>0.345</b>	± 0.069
MFI MRI *	%		<b>26.7</b>	<b>26.7</b>	± 8.0		<b>46.3</b>	<b>46.3</b>	± 10.0		<b>53.8</b>	<b>53.8</b>	± 12.0
GR RBC	10 <sup>6</sup> /µl		<b>3.39</b>	<b>3.39</b>	± 0.30		<b>3.38</b>	<b>3.38</b>	± 0.30		<b>3.40</b>	<b>3.40</b>	± 0.30
PIC RANGE	-		<b>27.0</b>	<b>27.0</b>	± 5.0		<b>27.0</b>	<b>27.0</b>	± 5.0		<b>27.0</b>	<b>27.0</b>	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.43 FRONT / RECTO 1300090305

# ABX Minotrol Retic



**LOT** RX 442  
Rev 1

**CONTROL**

(Exp.) 2023-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITS UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
XLR			XLR			XLR							
Méthode Manuelle / Manual RETIC	%	<b>1.9</b>			± 1.0	<b>5.7</b>			± 2.3	<b>11.9</b>			± 3.6
RETIC	%	<b>1.57</b>			± 0.79	<b>5.19</b>			± 1.30	<b>11.10</b>			± 2.22
RETIC	10 <sup>6</sup> /µl	<b>0.053</b>			± 0.027	<b>0.172</b>			± 0.043	<b>0.373</b>			± 0.075
MFI MRI *	%	<b>28.6</b>			± 8.0	<b>39.3</b>			± 10.0	<b>43.5</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.36</b>			± 0.30	<b>3.32</b>			± 0.30	<b>3.36</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.43 BACK / VERSO 1300090305

\* For research use only in Canada and USA