


LOT PX 442
Rev 1


CONTROL

 (Exp.) 2023-09-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0			± 0.4	8.6	8.6			± 1.0	18.0	18.0			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.35	2.35			± 0.16	4.66	4.66			± 0.20	5.25	5.25			± 0.25			
HB HGB	g/dl	6.1	6.1			± 0.4	14.0	14.0			± 0.5	17.1	17.1			± 0.6			
	g/l	61	61			± 4	140	140			± 5	171	171			± 6			
	mmol/l	3.79	3.79			± 0.25	8.69	8.69			± 0.31	10.62	10.62			± 0.37			
HT HCT	%	18.3	18.3			± 1.5	40.1	40.1			± 2.0	48.3	48.3			± 2.5			
	l/l	0.183	0.183			± 0.015	0.401	0.401			± 0.020	0.483	0.483			± 0.025			
VGM MCV	µm ³ fl	78	78			± 5	86	86			± 5	92	92			± 5			
TGMH MCH	pg	26.0	26.0			± 2.0	30.0	30.0			± 2.0	32.6	32.6			± 2.5			
	fmol	1.61	1.61			± 0.12	1.87	1.87			± 0.12	2.02	2.02			± 0.16			
CCMH MCHC	g/dl	33.3	33.3			± 3.0	34.9	34.9			± 3.0	35.4	35.4			± 3.0			
	g/l	333	333			± 30	349	349			± 30	354	354			± 30			
	mmol/l	20.67	20.67			± 1.86	21.69	21.69			± 1.86	21.99	21.99			± 1.86			
IDR RDW	%	15.0	15.0			± 4.0	12.5	12.5			± 4.0	12.0	12.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	66	66			± 20	250	250			± 30	466	466			± 50			
VPM MPV	µm ³ fl	8.9	8.9			± 2.0	9.4	9.4			± 2.0	9.3	9.3			± 2.0			
NEUT	#	1.51	1.51			± 0.35	4.63	4.63			± 0.90	13.25	13.25			± 1.90			
	%	50.2	50.2			± 10.0	53.8	53.8			± 10.0	73.6	73.6			± 10.0			
LYMPHO	#	1.17	1.17			± 0.33	3.27	3.27			± 0.70	2.90	2.90			± 1.50			
	%	39.0	39.0			± 12.0	38.0	38.0			± 8.0	16.1	16.1			± 8.0			
MONO	#	0.05	0.05			± 0.05	0.15	0.15			± 0.15	0.25	0.25			± 0.25			
	%	1.8	1.8			± 1.8	1.7	1.7			± 1.7	1.4	1.4			± 1.4			
EOS	#	0.19	0.19			± 0.19	0.31	0.31			± 0.31	0.86	0.86			± 0.86			
	%	6.2	6.2			± 6.2	3.6	3.6			± 3.6	4.8	4.8			± 4.8			
BASO	#	0.08	0.08			± 0.08	0.25	0.25			± 0.25	0.74	0.74			± 0.74			
	%	2.8	2.8			± 2.8	2.9	2.9			± 2.9	4.1	4.1			± 4.1			

LOT PX 442
Rev 1

CONTROL

 (Exp.) 2023-09-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0	3.0		± 0.4	8.6	8.6	8.4		± 1.0	18.4	18.1	17.8		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.41	2.38		± 0.16	4.66	4.67	4.62		± 0.20	5.23	5.21	5.22		± 0.25			
	g/dl	6.1	6.1	6.1		± 0.4	13.9	13.9	13.9		± 0.5	17.0	17.0	17.0		± 0.6			
HB HGB	g/l	61	61	61		± 4	139	139	139		± 5	170	170	170		± 6			
	mmol/l	3.79	3.79	3.79		± 0.25	8.63	8.63	8.63		± 0.31	10.56	10.56	10.56		± 0.37			
HT HCT	%	18.1	17.8	17.6		± 1.5	40.1	39.2	39.3		± 2.0	48.6	47.9	47.8		± 2.5			
	l/l	0.181	0.178	0.176		± 0.015	0.401	0.392	0.393		± 0.020	0.486	0.479	0.478		± 0.025			
VGM MCV	µm ³ ; fl	75	74	74.0		± 5	86	84	85.0		± 5	93	92	91.5		± 5			
TGMH MCH	pg	25.3	25.3	25.6		± 2.0	29.8	29.8	30.1		± 2.0	32.5	32.6	32.6		± 2.5			
	fmol	1.57	1.57	1.59		± 0.12	1.85	1.85	1.87		± 0.12	2.02	2.03	2.02		± 0.16			
	g/dl	33.7	34.2	34.6		± 3.0	34.7	35.4	35.4		± 3.0	35.0	35.5	35.6		± 3.0			
CCMH MCHC	g/l	337	342	346		± 30	347	354	354		± 30	350	355	356		± 30			
	mmol/l	20.96	21.24	21.51		± 1.86	21.54	22.00	21.98		± 1.86	21.70	22.02	22.10		± 1.86			
IDR RDW	%	15.5	15.5	14.0		± 4.0	12.5	12.5	11.5		± 4.0	12.0	12.0	11.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	67	67		± 20	258	257	241		± 30	476	479	447		± 50			
VPM MPV	µm ³ ; fl	8.7	8.8	7.9		± 2.0	8.5	8.6	8.3		± 2.0	8.4	8.6	8.1		± 2.0			
NEUT	#	1.47	1.43	1.44		± 0.35	4.58	4.62	4.54		± 0.90	13.41	13.30	13.35		± 1.90			
	%	49.0	47.6	48.0		± 10.0	53.2	53.7	54.0		± 10.0	72.9	73.5	75.0		± 10.0			
LYMPHO	#	1.20	1.24	1.23		± 0.33	3.35	3.34	3.23		± 0.70	3.05	2.97	2.85		± 1.50			
	%	40.0	41.2	41.0		± 12.0	38.9	38.8	38.5		± 8.0	16.6	16.4	16.0		± 8.0			
MONO	#	0.06	0.06	0.06		± 0.06	0.16	0.14	0.13		± 0.13	0.31	0.25	0.18		± 0.18			
	%	2.0	1.9	2.0		± 1.9	1.9	1.6	1.5		± 1.5	1.7	1.4	1.0		± 1.0			
EOS	#	0.19	0.20	0.18		± 0.18	0.27	0.26	0.25		± 0.25	0.86	0.83	0.71		± 0.71			
	%	6.2	6.5	6.0		± 6.0	3.1	3.0	3.0		± 3.0	4.7	4.6	4.0		± 4.0			
BASO	#	0.08	0.08	0.09		± 0.08	0.25	0.25	0.25		± 0.25	0.75	0.76	0.71		± 0.71			
	%	2.8	2.8	3.0		± 2.8	2.9	2.9	3.0		± 2.9	4.1	4.1	4.0		± 4.0			

Ref: TEMP-0821 Rev.48 FRONT / RECTO