

LOT

MX 442 L
Rev 1


CONTROL

L

 (Exp.)

2023-10-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE							
		MICROS															
		EMI CRP															
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l									2.0							± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l									2.27							± 0.15
HB HGB	g/dl									5.9							± 0.4
	g/l									59							± 4
	mmol/l									3.66							± 0.25
HT HCT	%									15.9							± 2.0
	l/l									0.159							± 0.020
VGM MCV	µm ³ ; fl									70							± 4
TGMH MCH	pg									26.0							± 2.0
	fmol									1.61							± 0.12
CCMH MCHC	g/dl									37.1							± 3.0
	g/l									371							± 30
	mmol/l									23.06							± 1.86
IDR RDW	%									14.0							± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l									72							± 20
VPM MPV	µm ³ ; fl									7.4							± 2.0
LYMPHO	#									1.23							± 0.40
	%									64.0							± 8.0
MONO	#									0.20							± 0.20
	%									7.5							± 6.0
GRANULO	#									0.57							± 0.40
	%									28.5							± 7.0

LOT		MX 442 N	CONTROL		N					 (Exp.)	2023-10-05 (YYYY-MM-DD)
		Rev 1	ABX Lysebio								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
				EMI CRP							
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			7.3							± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			4.58							± 0.18
HB HGB	g/dl			13.4							± 0.5
	g/l			134							± 5
	mmol/l			8.32							± 0.31
HT HCT	%			36.6							± 2.5
	l/l			0.366							± 0.025
VGM MCV	µm ³ ; fl			80							± 4
TGMH MCH	pg			29.3							± 2.0
	fmol			1.82							± 0.12
CCMH MCHC	g/dl			36.6							± 3.0
	g/l			366							± 30
	mmol/l			22.71							± 1.86
IDR RDW	%			14.0							± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			259							± 40
VPM MPV	µm ³ ; fl			8.0							± 2.0
LYMPHO	#			2.63							± 0.40
	%			36.0							± 5.0
MONO	#			0.55							± 0.30
	%			7.5							± 4.0
GRANULO	#			4.12							± 0.60
	%			56.5							± 6.0

Ret: TEMP-0830 Rev.42 BACK / VERSO 1300104144

LOT

MX 442 H
Rev 1

CONTROL

H

 (Exp.)

2023-10-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE							
		MICROS															
		EMI CRP															
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l									20.2							± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l									5.55							± 0.20
HB HGB	g/dl									17.7							± 0.6
	g/l									177							± 6
	mmol/l									10.99							± 0.37
HT HCT	%									48.8							± 3.0
	l/l									0.488							± 0.030
VGM MCV	µm ³ ; fl									88							± 4
TGMH MCH	pg									31.9							± 2.0
	fmol									1.98							± 0.12
CCMH MCHC	g/dl									36.2							± 3.0
	g/l									362							± 30
	mmol/l									22.51							± 1.86
IDR RDW	%									13.0							± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l									539							± 45
VPM MPV	µm ³ ; fl									8.0							± 2.0
LYMPHO	#									3.54							± 1.00
	%									17.5							± 5.0
MONO	#									1.21							± 0.80
	%									6.0							± 4.0
GRANULO	#									15.45							± 1.40
	%									76.5							± 7.0