


LOT PX 442

CONTROL

 (Exp.) **2023-09-05**
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE	
		CONTROL				L	CONTROL				N	CONTROL					H
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE
			DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0		± 0.4	8.8	8.8			± 1.0	18.6	18.6			± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.41		± 0.16	4.73	4.73			± 0.20	5.36	5.36			± 0.25		
HB HGB	g/dl	6.3	6.3		± 0.4	14.0	14.0			± 0.5	17.0	17.0			± 0.6		
	g/l	63	63		± 4	140	140			± 5	170	170			± 6		
	mmol/l	3.91	3.91		± 0.25	8.69	8.69			± 0.31	10.56	10.56			± 0.37		
HT HCT	%	19.0	19.0		± 1.5	41.2	41.2			± 2.0	49.3	49.3			± 2.5		
	l/l	0.190	0.190		± 0.015	0.412	0.412			± 0.020	0.493	0.493			± 0.025		
VGM MCV	µm ³ ; fl	79	79		± 5	87	87			± 5	92	92			± 5		
TGMH MCH	pg	26.1	26.1		± 2.0	29.6	29.6			± 2.0	31.7	31.7			± 2.5		
	fmol	1.62	1.62		± 0.12	1.84	1.84			± 0.12	1.97	1.97			± 0.16		
CCMH MCHC	g/dl	33.1	33.1		± 3.0	34.0	34.0			± 3.0	34.5	34.5			± 3.0		
	g/l	331	331		± 30	340	340			± 30	345	345			± 30		
	mmol/l	20.55	20.55		± 1.86	21.13	21.13			± 1.86	21.41	21.41			± 1.86		
IDR RDW	%	17.5	17.5		± 4.0	14.5	14.5			± 4.0	13.5	13.5			± 4.0		
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	69	69		± 20	260	260			± 30	477	477			± 50		
VPM MPV	µm ³ ; fl	9.1	9.1		± 2.0	9.1	9.1			± 2.0	9.1	9.1			± 2.0		
NEUT	#	1.64	1.64		± 0.35	5.00	5.00			± 0.90	13.90	13.90			± 1.90		
	%	54.5	54.5		± 10.0	56.8	56.8			± 10.0	74.6	74.6			± 10.0		
LYMPHO	#	0.91	0.91		± 0.33	2.90	2.90			± 0.70	2.64	2.64			± 1.50		
	%	30.4	30.4		± 12.0	33.0	33.0			± 8.0	14.2	14.2			± 8.0		
MONO	#	0.08	0.08		± 0.08	0.20	0.20			± 0.20	0.32	0.32			± 0.32		
	%	2.5	2.5		± 2.5	2.3	2.3			± 2.3	1.7	1.7			± 1.7		
EOS	#	0.22	0.22		± 0.22	0.30	0.30			± 0.30	0.87	0.87			± 0.87		
	%	7.4	7.4		± 7.4	3.4	3.4			± 3.4	4.7	4.7			± 4.7		
BASO	#	0.16	0.16		± 0.16	0.40	0.40			± 0.40	0.89	0.89			± 0.89		
	%	5.2	5.2		± 5.2	4.5	4.5			± 4.5	4.8	4.8			± 4.8		

Ref: TEMP-0821 Rev.48 BACK / VERSO