

# ABX Minotrol Retic



**LOT** RX 442  
Rev 1

**CONTROL**

(Exp.) 2023-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.9	1.9	± 1.0	5.7	5.7	± 2.3	11.9	11.9	± 3.6			
RETIC	%	1.29	1.29	± 0.65	4.98	4.98	± 1.25	10.16	10.16	± 2.03			
RETIC	10 <sup>6</sup> /µl	0.044	0.044	± 0.022	0.168	0.168	± 0.042	0.345	0.345	± 0.069			
MFI MRI *	%	26.7	26.7	± 8.0	46.3	46.3	± 10.0	53.8	53.8	± 12.0			
GR RBC	10 <sup>6</sup> /µl	3.39	3.39	± 0.30	3.38	3.38	± 0.30	3.40	3.40	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.43 FRONT / RECTO 1300090305

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 442  
Rev 1

**CONTROL**

(Exp.) 2023-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITS UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
XLR			XLR			XLR							
Méthode Manuelle / Manual RETIC	%	<b>1.9</b>			± 1.0	<b>5.7</b>			± 2.3	<b>11.9</b>			± 3.6
RETIC	%	<b>1.57</b>			± 0.79	<b>5.19</b>			± 1.30	<b>11.10</b>			± 2.22
RETIC	10 <sup>6</sup> /µl	<b>0.053</b>			± 0.027	<b>0.172</b>			± 0.043	<b>0.373</b>			± 0.075
MFI MRI *	%	<b>28.6</b>			± 8.0	<b>39.3</b>			± 10.0	<b>43.5</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.36</b>			± 0.30	<b>3.32</b>			± 0.30	<b>3.36</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.43 BACK / VERSO 1300090305

\* For research use only in Canada and USA