

ABX Minotrol CRP



LOT

MC 428

CONTROL



(Exp.) 2023-10-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio															
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA			
				EMI CRP	LC-767G	MS CRP	EMI CRP		LC-767G	MS CRP	EMI CRP	LC-767G		MS CRP					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0	2.1	± 0.4	7.5	7.4		7.8	± 1.0	19.9	20.2		20.8	± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.32	2.30	2.28	± 0.15	4.66	4.63	4.66	± 0.18	5.50	5.54	5.62	± 0.25						
HB HGB	g/dl	6.0	6.0	5.7	± 0.4	13.6	13.5	13.6	± 0.5	17.6	17.7	17.9	± 0.7						
	g/l	60	60	57	± 4	136	135	136	± 5	176	177	179	± 7						
	mmol/l	3.73	3.73	3.54	± 0.25	8.45	8.38	8.45	± 0.31	10.93	10.99	11.12	± 0.44						
HT HCT	%	16.2	16.3	16.0	± 2.0	37.3	37.0	36.8	± 2.5	49.0	48.8	48.8	± 3.0						
	l/l	0.162	0.163	0.160	± 0.020	0.373	0.370	0.368	± 0.025	0.490	0.488	0.488	± 0.030						
VGM MCV	µm ³ ; fl	70	71	70.0	± 4	80	80	79.0	± 4	89	88	86.0	± 4						
TGMH MCH	pg	25.9	26.1	25.0	± 2.4	29.2	29.2	29.2	± 2.4	32.0	31.9	31.9	± 2.8						
	fmol	1.61	1.62	1.55	± 0.15	1.81	1.81	1.81	± 0.15	1.99	1.98	1.98	± 0.18						
CCMH MCHC	g/dl	36.9	36.7	35.7	± 3.0	36.5	36.4	36.9	± 3.0	36.0	36.3	37.0	± 3.0						
	g/l	369	367	357	± 30	365	364	369	± 30	360	363	370	± 30						
	mmol/l	22.94	22.82	22.18	± 1.86	22.65	22.63	22.94	± 1.86	22.33	22.55	23.00	± 1.86						
IDR RDW	%	14.0	13.0	15.5	± 3.0	12.5	13.0	13.0	± 3.0	12.0	12.0	12.0	± 3.0						
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	78	75	68	± 20	262	256	259	± 40	532	538	554	± 60						
VPM MPV	µm ³ ; fl	7.6	8.5	8.9	± 2.0	8.1	8.2	8.7	± 2.0	8.0	8.1	8.4	± 2.0						
LYMPHO	#	1.3	1.3	N/A	± 0.2	2.7	2.6	N/A	± 0.5	3.5	3.4	N/A	± 1.1						
	%	64.0	63.5	N/A	± 8.0	36.0	35.0	N/A	± 6.0	17.5	17.0	N/A	± 5.0						
MONO	#	0.2	0.2	N/A	± 0.2	0.6	0.6	N/A	± 0.2	1.3	1.3	N/A	± 1.0						
	%	8.0	8.0	N/A	± 7.0	7.5	8.0	N/A	± 5.0	6.5	6.5	N/A	± 5.0						
GRANULO	#	0.6	0.6	N/A	± 0.2	4.2	4.2	N/A	± 0.5	15.1	15.5	N/A	± 1.5						
	%	28.0	28.5	N/A	± 7.0	56.5	57.0	N/A	± 6.0	76.0	76.5	N/A	± 7.0						
CRP	mg/dl	0.48	0.51	0.44	± 0.40	2.49	2.57	2.29	± 0.80	4.62	4.95	4.46	± 1.50						
	mg/l	4.8	5.1	4.4	± 4.0	24.9	25.7	22.9	± 8.0	46.2	49.5	44.6	± 15.0						

Ref: TEMP-1070 Rev.42 BACK/VERSO 1300092170