

LOT PX 442

Rev 1

CONTROL




(Exp.)

2023-09-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA					
			XL80	XLR					XL80	XLR					XL80	XLR				
GB	WBC	10 ³ /mm ³ ; 10 ⁹ /l	3.0	3.0				± 0.4	8.6	8.6				± 1.0	18.0	18.0				± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.35				± 0.16	4.66	4.66				± 0.20	5.25	5.25				± 0.25
HB	HGB	g/dl	6.1	6.1				± 0.4	14.0	14.0				± 0.5	17.1	17.1				± 0.6
		g/l	61	61				± 4	140	140				± 5	171	171				± 6
		mmol/l	3.79	3.79				± 0.25	8.69	8.69				± 0.31	10.62	10.62				± 0.37
HT	HCT	%	18.3	18.3				± 1.5	40.1	40.1				± 2.0	48.3	48.3				± 2.5
		l/l	0.183	0.183				± 0.015	0.401	0.401				± 0.020	0.483	0.483				± 0.025
VGM	MCV	µm ³ ·fl	78	78				± 5	86	86				± 5	92	92				± 5
TGMH	MCH	pg	26.0	26.0				± 2.0	30.0	30.0				± 2.0	32.6	32.6				± 2.5
		fmol	1.61	1.61				± 0.12	1.87	1.87				± 0.12	2.02	2.02				± 0.16
CCMH	MCHC	g/dl	33.3	33.3				± 3.0	34.9	34.9				± 3.0	35.4	35.4				± 3.0
		g/l	333	333				± 30	349	349				± 30	354	354				± 30
		mmol/l	20.67	20.67				± 1.86	21.69	21.69				± 1.86	21.99	21.99				± 1.86
IDR	RDW	%	15.0	15.0				± 4.0	12.5	12.5				± 4.0	12.0	12.0				± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	66	66				± 20	250	250				± 30	466	466				± 50
VPM	MPV	µm ³ ·fl	8.9	8.9				± 2.0	9.4	9.4				± 2.0	9.3	9.3				± 2.0
NEUT	#		1.51	1.51				± 0.35	4.63	4.63				± 0.90	13.25	13.25				± 1.90
		%	50.2	50.2				± 10.0	53.8	53.8				± 10.0	73.6	73.6				± 10.0
LYMPHO	#		1.17	1.17				± 0.33	3.27	3.27				± 0.70	2.90	2.90				± 1.50
		%	39.0	39.0				± 12.0	38.0	38.0				± 8.0	16.1	16.1				± 8.0
MONO	#		0.05	0.05				± 0.05	0.15	0.15				± 0.15	0.25	0.25				± 0.25
		%	1.8	1.8				± 1.8	1.7	1.7				± 1.7	1.4	1.4				± 1.4
EOS	#		0.19	0.19				± 0.19	0.31	0.31				± 0.31	0.86	0.86				± 0.86
		%	6.2	6.2				± 6.2	3.6	3.6				± 3.6	4.8	4.8				± 4.8
BASO	#		0.08	0.08				± 0.08	0.25	0.25				± 0.25	0.74	0.74				± 0.74
		%	2.8	2.8				± 2.8	2.9	2.9				± 2.9	4.1	4.1				± 4.1

 (Exp.) **2023-09-05**
 (YYYY - MM - DD)

LOT

PX 442

Rev 1

CONTROL

(Exp.)

2023-09-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCRP				60 60C+ ES60	MS60	MSCRP				60 60C+ ES60	MS60	MSCRP			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0	3.0			± 0.4	8.6	8.6	8.4			± 1.0	18.4	18.1	17.8			± 2.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.41	2.38			± 0.16	4.66	4.67	4.62			± 0.20	5.23	5.21	5.22			± 0.25
HB HGB	g/dl	6.1	6.1	6.1			± 0.4	13.9	13.9	13.9			± 0.5	17.0	17.0	17.0			± 0.6
	g/l	61	61	61			± 4	139	139	139			± 5	170	170	170			± 6
	mmol/l	3.79	3.79	3.79			± 0.25	8.63	8.63	8.63			± 0.31	10.56	10.56	10.56			± 0.37
HT HCT	%	18.1	17.8	17.6			± 1.5	40.1	39.2	39.3			± 2.0	48.6	47.9	47.8			± 2.5
	l/l	0.181	0.178	0.176			± 0.015	0.401	0.392	0.393			± 0.020	0.486	0.479	0.478			± 0.025
VGM MCV	µm ³ ·fl	75	74	74.0			± 5	86	84	85.0			± 5	93	92	91.5			± 5
TGMH MCH	pg	25.3	25.3	25.6			± 2.0	29.8	29.8	30.1			± 2.0	32.5	32.6	32.6			± 2.5
	fmol	1.57	1.57	1.59			± 0.12	1.85	1.85	1.87			± 0.12	2.02	2.03	2.02			± 0.16
CCMH MCHC	g/dl	33.7	34.2	34.6			± 3.0	34.7	35.4	35.4			± 3.0	35.0	35.5	35.6			± 3.0
	g/l	337	342	346			± 30	347	354	354			± 30	350	355	356			± 30
	mmol/l	20.96	21.24	21.51			± 1.86	21.54	22.00	21.98			± 1.86	21.70	22.02	22.10			± 1.86
IDR RDW	%	15.5	15.5	14.0			± 4.0	12.5	12.5	11.5			± 4.0	12.0	12.0	11.0			± 4.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	67	67			± 20	258	257	241			± 30	476	479	447			± 50
VPM MPV	µm ³ ·fl	8.7	8.8	7.9			± 2.0	8.5	8.6	8.3			± 2.0	8.4	8.6	8.1			± 2.0
NEUT	#	1.47	1.43	1.44			± 0.35	4.58	4.62	4.54			± 0.90	13.41	13.30	13.35			± 1.90
	%	49.0	47.6	48.0			± 10.0	53.2	53.7	54.0			± 10.0	72.9	73.5	75.0			± 10.0
LYMPHO	#	1.20	1.24	1.23			± 0.33	3.35	3.34	3.23			± 0.70	3.05	2.97	2.85			± 1.50
	%	40.0	41.2	41.0			± 12.0	38.9	38.8	38.5			± 8.0	16.6	16.4	16.0			± 8.0
MONO	#	0.06	0.06	0.06			± 0.06	0.16	0.14	0.13			± 0.13	0.31	0.25	0.18			± 0.18
	%	2.0	1.9	2.0			± 1.9	1.9	1.6	1.5			± 1.5	1.7	1.4	1.0			± 1.0
EOS	#	0.19	0.20	0.18			± 0.18	0.27	0.26	0.25			± 0.25	0.86	0.83	0.71			± 0.71
	%	6.2	6.5	6.0			± 6.0	3.1	3.0	3.0			± 3.0	4.7	4.6	4.0			± 4.0
BASO	#	0.08	0.08	0.09			± 0.08	0.25	0.25	0.25			± 0.25	0.75	0.76	0.71			± 0.71
	%	2.8	2.8	3.0			± 2.8	2.9	2.9	3.0			± 2.9	4.1	4.1	4.0			± 4.0

Ref: TEMP-0821 Rev.48 FRONT / RECTO