

ABX Minocal


 (Exp.) 2023-09-05
 (YYYY - MM - DD)

LOT		CX 485		CAL		(Exp.) 2023-09-05 (YYYY - MM - DD)					
		Rev 1									
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60	DX120 - DF120	60 60 C+ ES60	XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.1	9.2	9.3	9.2	9.4	9.2	9.4	9.3	9.2	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.54	4.54	4.54	4.55	4.58	4.52	4.58	4.57	4.55	± 0.06
HB HGB	g/dl	13.3	13.3	13.5	13.4	13.4	13.3	13.4	13.5	13.4	± 0.2
	g/l	133	133	135	134	134	133	134	135	134	± 2
	mmol/l	8.26	8.26	8.38	8.32	8.32	8.26	8.32	8.38	8.32	± 0.12
HT HCT	%	37.9	37.5	37.0	36.9	36.6	36.2	36.6	37.2	36.9	± 1.0
	l/l	0.379	0.375	0.370	0.369	0.366	0.362	0.366	0.372	0.369	± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	235	242	230	231	244	239	244	238	231	± 10
VMP MPV	µm ³ ; fl	8.5	8.6	N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.49 BACK / VERSO 1300090308