

Yumizen G DDi 2 for YUMIZEN G1500/G1550



Reference No.:		1300036391	
Lot No.:		930501 (rev. 1)	
Expiry date:		2024-05-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G1500/G1550	µg(FEU)/mL	ΔOD
		4,00	0,2945
		3,00	0,2290
		2,00	0,1405
		1,00	0,0595
		0,50	0,0280
		0,25	0,0145
Slope value		0,07	

Yumizen G DDi 2 for YUMIZEN G800



Reference No.:		1300036391	
Lot No.:		930501 (rev. 1)	
Expiry date:		2024-05-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G800	µg(FEU)/mL	ΔOD
		4,00	0,2750
		3,00	0,2200
		2,00	0,1405
		1,00	0,0580
		0,50	0,0260
		0,25	0,0125
Slope value		0,06	



Yumizen G DDi 2 for YUMIZEN G400 DDi/G200

Reference No.:		1300036391	
Lot No.:		930501 (rev. 1)	
Expiry date:		2024-05-31	
Coagulometers		MASTER CURVE	
		µg(FEU)/mL	ΔOD
Optical	Yumizen G400 DDi/G200	4,00	0,1130
		3,00	0,0850
		2,00	0,0545
		1,00	0,0257
		0,50	0,0128
		0,25	0,0068
		Slope value	0,02
Important notice			
<p>For Yumizen G400 DDi and Yumizen G200, make sure that the end time is set to 180 seconds in the D-Dimer setup and the curve type Lin-Lin p-p is selected in Calibration > D-Dimer > ug FEU/ml. For more information, please contact your local HORIBA Medical representative.</p>			