



Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 3 0 4 1 2 4 6 4 1 2 0 4 2 5 D



D 5 1 1 7 1 5 8 0 5 9 0 9 9 0 8 9 D



D 6 1 3 4 1 1 4 1 5 5 0 6 0 1 0 0 D



D 7 0 9 0 1 3 4 0 9 1 1 2 3 0 7 4 D



D 8 1 2 3 0 8 1 1 2 1 3 4 0 4 6 0 D



D 9 2 9 2 3 9 5 2 0 5 3 0 7 2 1 0 D



D 0 3 1 5 3 2 9 D

Reference No.:		1300036412	
Lot No:		930412 (rev. 2)	
Expiry date:		2025-04-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	11,7-15,8
	Prothrombin %	%	59-99
	Prothrombin INR	INR	0,89-1,34
PT Liq (*)	Prothrombin Time	sec	11,4-15,5 (*)
	Prothrombin %	%	60-100 (*)
	Prothrombin INR	INR	0,90-1,34 (*)
PT R	Prothrombin Time	sec	9,1-12,3
	Prothrombin %	%	74-123
	Prothrombin INR	INR	0,81-1,21
APTT	Partial Thromboplastin Time	sec	34,0-46,0
APTT Liq	Partial Thromboplastin Time	sec	29,2-39,5
FIB Cl	Fibrinogen	g/L	2,05-3,07
TT	Thrombin Time	sec	21,0-31,5
AT	Antithrombin	%	92-125

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.

(*) Warning! This range has to be used only with Yumizen G PT Liq 4 lots : 920315 and higher except for the Yumizen G PT Liq 4 lots 920611 & 920705. For the other lots, including 920611 & 920705, please refer to the previous revision of values sheet.



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 3 0 4 1 2 5 6 4 1 2 0 4 2 5 D



D 5 1 6 2 2 4 3 0 2 8 0 4 7 1 4 1 D



D 6 2 1 2 1 7 2 5 8 0 2 6 0 4 4 D



D 7 1 5 7 2 3 6 1 3 0 1 9 5 0 3 5 D



D 8 0 5 9 1 2 2 1 8 3 6 5 3 8 8 3 D



D 9 6 5 5 8 8 6 0 6 9 1 6 0 0 0 D



D 0 9 9 9 3 8 0 D

Reference No.:		1300036412	
Lot No.:		930412 (rev. 2)	
Expiry date:		2025-04-30	
Optical Kit	Yumizen G Line Parameter	Unit	Range
PT	Prothrombin Time	sec	16,2-24,3
	Prothrombin %	%	28-47
	Prothrombin INR	INR	1,41-2,12
PT Liq (*)	Prothrombin Time	sec	17,2-25,8 (*)
	Prothrombin %	%	26-44 (*)
	Prothrombin INR	INR	1,57-2,36 (*)
PT R	Prothrombin Time	sec	13,0-19,5
	Prothrombin %	%	35-59
	Prothrombin INR	INR	1,22-1,83
APTT	Partial Thromboplastin Time	sec	65,3-88,3
APTT Liq	Partial Thromboplastin Time	sec	65,5-88,6
FIB Cl	Fibrinogen	g/L	0,69-1,60
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	38-64

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.

(*) Warning! This range has to be used only with Yumizen G PT Liq 4 lots : 920315 and higher except for the Yumizen G PT Liq 4 lots 920611 & 920705. For the other lots, including 920611 & 920705, please refer to the previous revision of values sheet.