

# Yumizen G PT Reco 5 for Yumizen G Line



|  |                       |                        |                     |             |
|--|-----------------------|------------------------|---------------------|-------------|
| <b>Reference No.:</b>  |                       | <b>1300036375</b>      |                     |             |
| <b>Lot No.:</b>  |                       | <b>930413 (rev. 1)</b> |                     |             |
| <b>Expiry date:</b>  |                       | <b>2025-04-30</b>      |                     |             |
| <b>Coagulometers</b>   |                       | <b>ISI</b>             | <b>MASTER CURVE</b> |             |
| <b>Optical</b>   | <b>Yumizen G Line</b> |                        | <b>%</b>            | <b>SEC</b>  |
|  |                       | <b>100</b>             | <b>10,2</b>         |             |
|  |                       | <b>50</b>              | <b>15,0</b>         |             |
|  |                       | <b>33</b>              | <b>19,9</b>         |             |
|  |                       | <b>25</b>              | <b>24,5</b>         |             |
|  |                       | <b>1,00</b>            | <b>MNPT</b>         | <b>10,2</b> |
| <b>Important notice</b>  |                       |                        |                     |             |
| <p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>For Yumizen G200, Yumizen G400 and Yumizen G400 DDi, the <b>Min. Time</b> must be set to 7 seconds and the <b>Lag time</b> to 5 seconds in the PT setup. Please refer to the Yumizen G PT Reco 5 reagent notice.</p> |                       |                        |                     |             |