

ABX Erytrol



LOT

NX 443

Rev 1

CONTROL



(Exp.) 2023-11-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA	DX120		DX NEXUS	DX120	DX NEXUS		DX120	DX NEXUS		
		DX120	DX NEXUS										
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	7.5	7.5		± 0.8	18.1	18.1		± 1.5	2.8	2.8		± 0.4
GBC CWBC	10 ³ /mm ³ ; 10 ⁹ /l	7.3	7.3		± 0.8	16.6	16.6		± 1.4	1.6	1.6		± 0.4
ERB	%	3.1	3.1		± 1.7	8.9	8.9		± 3.3	72.6	72.6		± 50.0
ERB	#	0.226	0.226		± 0.100	1.477	1.477		± 0.700	1.162	1.162		± 0.500

1219930076-C

Ref: TEMP-0822 Rev.36

GBC / CWBC

* les valeurs sont données à titre d'indication et doivent être utilisées comme RuO (Utilisation en Recherche Seulement) Indicative expected values, must be used as RuO (Research used only)